

QIBA CT Small Lung Nodule Profile

CTLX1 Validation Study

July 12, 2021 to October ?, 2021


Goal: Verify that the CTLX1 phantom analysis values are correct

QIBA SLN CT Image Quality Metrics

1. Edge Enhancement
2. 3D Resolution
3. Resolution Aspect
4. HU Bias (Air & Acrylic)
5. Image Noise (Air & Acrylic)
6. Spatial Warping

Distance From Iso-Center

1. 0 mm
2. 100 mm
3. 200 mm (no other phantom can do this)



CTLX1 Phantom

VolView - 1.3.12.2.1107...814800017479 : CTLX 2 CTLX 2 PHANTOM : 06282021 : 1 : 9 : 630 : CT

File Edit View Window Help

Color/Opacity Presets

Ima...	Type	Comment
	CT	AAA
	CT	AAA2
	CT	Bone
	CT	Cardiac
	CT	Cardiac2
	CT	Cardiac3
	CT	Chest Contrast Enhanced
	CT	Chest Vessels

Color/Opacity Settings

Component: 1

Enable Shading

Scalar Opacity Mapping:
[-1014.18, 1257.79]

Scalar Color Mapping:
[-1014.18, 1257.79]

Component Weight(s):
1: 1

Volume Rendering

Volume
CT SIEMENS SOMATOM Definition AS CTAWP64263
Exam: 1
Series: 9 (B40 0.6 Axials)
1.3.12.2.11...14800017479
MT SINAI RAD ASSOCIATES x34962
CTLX 2 CTLX 2 PHANTOM
ID: 06282021
50 year(s) O 11/10/70
06/28/21 15:39:58

mA: 72
kVp: 120
Thick: 0.6 mm
Kernel: B40f
VolView 3.4

Axial
CT SIEMENS SOMATOM Definition AS CTAWP64263
Exam: 1
Series: 9 (B40 0.6 Axials)
Image: 91 / 255
54 mm
A
R
L
P
1.3.12.2.11...14800017479
MT SINAI RAD ASSOCIATES x34962
CTLX 2 CTLX 2 PHANTOM
ID: 06282021
50 year(s) O 11/10/70
06/28/21 15:39:58

mA: 72
kVp: 120
Thick: 0.6 mm
Kernel: B40f
VolView 3.4

WW/WL: 2587 / -27

Coronal
CT SIEMENS SOMATOM Definition AS CTAWP64263
Exam: 1
Series: 9 (B40 0.6 Axials)
Image: 329 / 512
-155.699 mm
S
R
L
I
1.3.12.2.11...14800017479
MT SINAI RAD ASSOCIATES x34962
CTLX 2 CTLX 2 PHANTOM
ID: 06282021
50 year(s) O 11/10/70
06/28/21 15:39:58

mA: 72
kVp: 120
Thick: 0.6 mm
Kernel: B40f
VolView 3.4

WW/WL: 2587 / -27


Sagittal
CT SIEMENS SOMATOM Definition AS CTAWP64263
Exam: 1
Series: 9 (B40 0.6 Axials)
Image: 253 / 512
-3.41797 mm
S
A
P
I
1.3.12.2.11...14800017479
MT SINAI RAD ASSOCIATES x34962
CTLX 2 CTLX 2 PHANTOM
ID: 06282021
50 year(s) O 11/10/70
06/28/21 15:39:58

mA: 72
kVp: 120
Thick: 0.6 mm
Kernel: B40f
VolView 3.4

WW/WL: 2587 / -27

12:13 PM

CTLX1 Phantom



**Slice Selection
Is A Source Of
Variability
For Manual
Measurement**

**Non-alignment
& Large FOV Is
Supported by
CTLX1 Analysis**

Goal: Verify that the CTLX1 phantom analysis values are correct

- Validation Study
 - The CTLX1 was CT scanned on a Siemens SOMATOM Definition AS varying:
 - Slice Thickness : 0.6 mm and 1.0 mm.
 - Reconstruction Kernel : B30f, B40f, and B60f (highly edge enhancing).
 - A CTLX1 scan from a 16 slice model known to exhibit spatial warping was obtained.
 - Accumetra's automated software was run on the main 6 scans and results distributed. Automated analysis of the scan known to exhibit spatial warping was done later due to an oversight. Accumetra in plane PSF sigma was converted to an MTF 50 for comparison.
 - OHSU and Mount Sinai received the 7 scans and manually measured the QIBA SLN metrics used by the QIBA SLN Profile, the most challenging being in-plane spatial resolution.

Measurement Methods


For each of three modules in the CTLX1:

1. Edge Enhancement:

- Pick a slice and manually measure.

2. 3D Resolution (in-plane):

- Pick a slice and manually measure MTF 50.
- Use a “standard” edge method to measure in-plane MTF 50.
 - Suggested use of AAPM TG 233 report – although not enough detail for perfect reproducibility.
- Verify expected behavior is present:
 - B30f in-plane resolution is lower than B40f for all datasets and modules.
 - Slice thickness does not significantly change resolution values.
 - Lower resolution is observed as a function of distance from iso-center.



Measurement Methods

For each of three modules in the CTLX1:







2. Z Resolution (slice thickness)
 - Report the requested slice thickness (passed).
3. Resolution Aspect
 - Measurement not needed since aspect ratio is based on resolution measures.
4. HU Bias
 - Measure the mean HU value within the air and acrylic cylinder.
5. Image Noise (HU SD)
 - Measure the HU SD noise level within the air and acrylic cylinder.
6. Spatial Warping
 - Visually grade whether spatial warping is present on a coronal view of the CTLX1 Delrin cylinder.

CTLX1 Scans

Scan	Model	Kernel	Slice Thickness	Pass/Fail & Reason
1	SOMATOM Def AS	B30f	0.6 mm	PASS
2	SOMATOM Def AS	B30f	1.0 mm	PASS
3	SOMATOM Def AS	B40f	0.6 mm	PASS
4	SOMATOM Def AS	B40f	1.0 mm	PASS
5	SOMATOM Def AS	B60f	0.6 mm	FAIL – EE
6	SOMATOM Def AS	B60f	1.0 mm	FAIL - EE
7	LightSpeed16	LUNG	1.25 mm	FAIL – SW, EE

Multi-center Study Results





OHSU, Mount Sinai, and Accumetra : Edge Enhancement







Edge Enhancement







OHSU, Mount Sinai, and Accumetra : MTF 50







OHSU, Mount Sinai, and Accumetra : Air HU







OHSU, Mount Sinai, and Accumetra : Air HU SD



OHSU, Mount Sinai, and Accumetra : Acrylic HU



OHSU, Mount Sinai, and Accumetra : Acrylic HU SD



OHSU Results : Spatial Warping

Z Spatial Warping							
CT Scan	OHSU Spatial Warping Y/N			Accumetra Spatial Warping Y/N			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
	1 CTLX1-B30-0.6	N	N	N	0.046	0.04	0.05
2 CTLX1-B30-1.0	N	N	N	0.07	0.055	0.059	
3 CTLX1-B40-0.6	N	N	N	0.045	0.04	0.041	
4 CTLX1-B40-1.0	N	N	N	0.078	0.056	0.061	
5 CTLX1-B60-0.6	N	N	N	0.081	0.073	0.07	
6 CTLX1-B60-1.0	N	N	N	0.106	0.099	0.083	
7 ExtraScan	N	N	Y	0.12	0.284	0.432	

Any value > 0.3 is considered positive for Z spatial warping

100% agreement on spatial warping (but no visual cutoff guidance was provided)


Mount Sinai Results : Spatial Warping

Z Spatial Warping							
		Mount Sinai			Accumetra		
		Air HU SD			Spatial Warping Y/N		
CT Scan	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
1 CTLX1-B30-0.6	No	No	No	0.046	0.04	0.05	
2 CTLX1-B30-1.0	No	No	No	0.07	0.055	0.059	
3 CTLX1-B40-0.6	No	No	No	0.045	0.04	0.041	
4 CTLX1-B40-1.0	No	No	No	0.078	0.056	0.061	
5 CTLX1-B60-0.6	No	No	No	0.081	0.073	0.07	
6 CTLX1-B60-1.0	No	No	No	0.106	0.099	0.083	
7 ExtraScan	Yes	Yes	Yes	0.12	0.284	0.432	

Any value > 0.3 is considered positive for Z spatial warping

90.5% agreement on spatial warping (but no visual cutoff guidance was provided)

Spatial Warping



Many, Many Thanks

- Mount Sinai
 - Raj Subramaniam
 - Shirong Zhang
 - Jo-Ann Provencher
- OHSU
 - Tom Griglock
 - Celeste Leary
 - Grace Eliason

Thank You

Mount Sinai Results



Mount Sinai Results : Edge Enhancement

- 0.041

Edge Enhancement							
CT Scan	Mount Sinai			Accumetra			
	Edge Enhancement			Edge Enhancement			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
1 CTLX1-B30-0.6	1.030828959	1.016215453	1.014038764	1.039	1.044	1.055	
2 CTLX1-B30-1.0	1.028272744	1.02682247	1.017312554	1.038	1.043	1.05	
3 CTLX1-B40-0.6	1.023124604	1.018682454	1.018179733	1.042	1.039	1.04	
4 CTLX1-B40-1.0	1.021920053	1.007508767	1.005475721	1.04	1.041	1.046	
5 CTLX1-B60-0.6	1.104083428	1.064212223	1.044380626	1.097	1.098	1.1	
6 CTLX1-B60-1.0	1.112645846	1.059752188	1.078907765	1.094	1.104	1.105	
7 ExtraScan				1.192	1.205	1.185	

83% (5/6) agreement on Failure Due To Edge Enhancement (EE <= 1.05)

Mount Sinai Results : Edge Enhancement



Mount Sinai Results : Edge Enhancement





Mount Sinai Results : In Plane Resolution

In-Plane Resolution		Mount Sinai			Accumetra		
		MTF50 (lp/mm)			PSF to MTF50 (lp/mm)		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	0.2980	0.2944	0.2770	0.336183611	0.301375362	0.225969709
2	CTLX1-B30-1.0	0.2920	0.2928	0.2610	0.342452641	0.313238986	0.229662717
3	CTLX1-B40-0.6	0.3055	0.3091	0.2688	0.406668244	0.361627107	0.259289129
4	CTLX1-B40-1.0	0.3141	0.3114	0.2709	0.408383676	0.374706917	0.270135144
5	CTLX1-B60-0.6	0.3606	0.3866	0.3889	0.817618307	0.817618307	0.817618307
6	CTLX1-B60-1.0	0.3699	0.4057	0.3870	0.817618307	0.817618307	0.817618307
7	ExtraScan				0.817618307	0.817618307	0.610932006

- ✓ Good levels of agreement on in-plane resolution
- ✓ B30f resolution is lower than B40f.
- ✓ Slice thickness does not significantly change resolution values.
- ✓ Lower resolution is observed as a function of distance from iso-center.

Mount Sinai Results : In-Plane Resolution

IEC Longitudinal Drift < 0.75 mm



Mount Sinai
Accumetra

We expect differences due to high edge enhancement (x/y resolution not valid when EE is high) and warping.


Mount Sinai Results : Air & Acrylic HU Bias

Air HU							
		Mount Sinai			Accumetra		
		Air Mean HU			Air Mean HU		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	-999.5	-997.1	-993.2	-997.460	-996.610	-992.810
2	CTLX1-B30-1.0	-997.8	-997.8	-992.2	-997.77	-996.580	-992.820
3	CTLX1-B40-0.6	-997.8	-998.2	-991.1	-996.6	-996.29	-990.870
4	CTLX1-B40-1.0	-997.6	-996.2	-991.0	-996.71	-995.92	-990.78
5	CTLX1-B60-0.6	-989.9	-986.1	-978.0	-990.48	-986.51	-975.17
6	CTLX1-B60-1.0	-988.9	-986.6	-975.3	-990.82	-986.28	-974.59
7	ExtraScan	-998.4	-994.9	-980.6	-998.79	-993.54	-987.36
Acrylic HU							
		Mount Sinai			Accumetra		
		Acrylic Mean HU			Acrylic Mean HU		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	135.6	138.9	154.0	132.71	132.4	143.27
2	CTLX1-B30-1.0	136.5	141.6	157.6	132.09	132.2	143.72
3	CTLX1-B40-0.6	132.7	135.8	149.0	128.09	128.19	139.83
4	CTLX1-B40-1.0	137.9	142.4	149.4	127.85	128.12	139.37
5	CTLX1-B60-0.6	115.7	104.4	108.8	110.76	107.51	117.1
6	CTLX1-B60-1.0	112.2	112.7	118.0	111.05	109.56	120.85
7	ExtraScan	143.3	139.9	147.8	139.22	149.46	154.23

-2.04

14.28

Mount Sinai Results : Air & Acrylic Bias



Mount Sinai Results : Air & Acrylic Image Noise (HU SD)

Air HU SD							
CT Scan	Mount Sinai Air HU SD			Accumetra Air HU SD			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
	1 CTLX1-B30-0.6	8.0	12.8	17.6	8.910	11.33	
2 CTLX1-B30-1.0	5.6	9.0	12.4	6.1	7.41	9.98	
3 CTLX1-B40-0.6	10.7	12.4	16.1	9.6	13.15	17.41	
4 CTLX1-B40-1.0	7.9	7.8	10.5	6.25	8.31	10.97	
5 CTLX1-B60-0.6	30.6	42.4	52.2	29.85	41.92	57.82	
6 CTLX1-B60-1.0	19.8	26.8	37.6	18.72	27.36	40.5	
7 ExtraScan	22.1	28.1	41.0	21.34	28.22	35.79	
Acrylic HU SD							
CT Scan	Mount Sinai Air HU SD			Accumetra Acrylic HU SD			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
	1 CTLX1-B30-0.6	8.0	16.9	18.1	10.06	13.59	
2 CTLX1-B30-1.0	10.2	13.6	13.8	6.5	9.55	10.92	
3 CTLX1-B40-0.6	16.9	18.5	18.8	11.59	15.35	18.91	
4 CTLX1-B40-1.0	15.3	19.0	19.8	7.51	9.76	11.88	
5 CTLX1-B60-0.6	43.6	68.1	79.6	43.98	69.29	83.69	
6 CTLX1-B60-1.0	28.7	46.4	47.1	27.82	44.42	52.32	
7 ExtraScan	32.0	36.0	53.5	28.71	36.59	52.72	

Mount Sinai Results : Air & Acrylic Image Noise (HU SD)

B30f 0.6mm

B30f 1.0mm


B40f 0.6mm

B40f 1.0mm

B60f 0.6mm

B60f 1.0mm

LUNG 1.25mm Warped



Mount Sinai
Accumetra

We expect differences due to high edge enhancement (noise impacted when EE is high) and warping.


Mount Sinai Results : Spatial Warping

Z Spatial Warping							
		Mount Sinai			Accumetra		
		Air HU SD			Spatial Warping Y/N		
CT Scan	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
1 CTLX1-B30-0.6	No	No	No	0.046	0.04	0.05	
2 CTLX1-B30-1.0	No	No	No	0.07	0.055	0.059	
3 CTLX1-B40-0.6	No	No	No	0.045	0.04	0.041	
4 CTLX1-B40-1.0	No	No	No	0.078	0.056	0.061	
5 CTLX1-B60-0.6	No	No	No	0.081	0.073	0.07	
6 CTLX1-B60-1.0	No	No	No	0.106	0.099	0.083	
7 ExtraScan	Yes	Yes	Yes	0.12	0.284	0.432	

Any value > 0.3 is considered positive for Z spatial warping

90.5% agreement on spatial warping (but no visual cutoff guidance was provided)

Mount Sinai Results : Spatial Warping



Preliminary Findings & Observations

- From OHSU Review on 10/1/21
 - Remeasure using multiple slices.
 - We should put SD on graphs (particularly acrylic).

OHSU Results



OHSU Results : Edge Enhancement

- 0.02

Edge Enhancement		OHSU			Accumetra		
		Edge Enhancement			Edge Enhancement		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	1.038	1.036	1.035	1.039	1.044	1.055
2	CTLX1-B30-1.0	1.034	1.033	1.039	1.038	1.043	1.05
3	CTLX1-B40-0.6	1.037	1.019	1.039	1.042	1.039	1.04
4	CTLX1-B40-1.0	1.037	1.034	1.041	1.04	1.041	1.046
5	CTLX1-B60-0.6	1.125	1.110	1.079	1.097	1.098	1.1
6	CTLX1-B60-1.0	1.124	1.138	1.085	1.094	1.104	1.105
7	ExtraScan	1.152	1.193	1.134	1.192	1.205	1.185

100% agreement on Failure Due To Edge Enhancement (EE <= 1.05)

OHSU Results : Edge Enhancement



OHSU
Accumetra

OHSU Results : Edge Enhancement



OHSU Results : In Plane Resolution



In-Plane Resolution		OHSU			Accumetra		
		MTF50 (lp/mm)			PSF to MTF50 (lp/mm)		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	0.351	0.325	0.279	0.336183611	0.301375362	0.225969709
2	CTLX1-B30-1.0	0.349	0.323	0.282	0.342452641	0.313238986	0.229662717
3	CTLX1-B40-0.6	0.388	0.340	0.298	0.406668244	0.361627107	0.259289129
4	CTLX1-B40-1.0	0.385	0.346	0.301	0.408383676	0.374706917	0.270135144
5	CTLX1-B60-0.6	0.651	0.551	0.472	0.817618307	0.817618307	0.817618307
6	CTLX1-B60-1.0	0.687	0.533	0.476	0.817618307	0.817618307	0.817618307
7	ExtraScan	0.724	0.443	0.396	0.817618307	0.817618307	0.610932006

- ✓ High levels of agreement on in-plane resolution
- ✓ B30f resolution is lower than B40f.
- ✓ Slice thickness does not significantly change resolution values.
- ✓ Lower resolution is observed as a function of distance from iso-center.

0.053

OHSU Results : In-Plane Resolution

IEC Longitudinal
Drift < 0.75 mm



These Scans All Failed EE

OHSU
Accumetra

We expect differences due to high edge enhancement (x/y resolution not valid when EE is high) and warping.


OHSU Results : Air & Acrylic HU Bias

Air HU		OHSU			Accumetra		
		Air Mean HU			Air Mean HU		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	-1000.525	-999.548	-993.452	-997.460	-996.610	-992.810
2	CTLX1-B30-1.0	-1000.389	-1000.096	-992.516	-997.77	-996.580	-992.820
3	CTLX1-B40-0.6	-1001.006	-994.625	-994.629	-996.6	-996.29	-990.870
4	CTLX1-B40-1.0	-998.698	-997.144	-993.035	-996.71	-995.92	-990.78
5	CTLX1-B60-0.6	-989.127	-987.739	-974.875	-990.48	-986.51	-975.17
6	CTLX1-B60-1.0	-991.645	-983.287	-970.175	-990.82	-986.28	-974.59
7	ExtraScan	-1000.309	-999.02	-987.47	-998.79	-993.54	-987.36
Acrylic HU		OHSU			Accumetra		
		Acrylic Mean HU			Acrylic Mean HU		
CT Scan		Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm
1	CTLX1-B30-0.6	140.349	143.405	151.621	132.71	132.4	143.27
2	CTLX1-B30-1.0	137.878	145.03	153.219	132.09	132.2	143.72
3	CTLX1-B40-0.6	135.379	140.884	142.054	128.09	128.19	139.83
4	CTLX1-B40-1.0	134.813	139.952	146.407	127.85	128.12	139.37
5	CTLX1-B60-0.6	113.698	116.406	122.362	110.76	107.51	117.1
6	CTLX1-B60-1.0	112.13	116.808	126.714	111.05	109.56	120.85
7	ExtraScan	153.347	151.873	142.828	139.22	149.46	154.23

-4.406

12.83

OHSU Results : Air & Acrylic Bias



OHSU Results : Air & Acrylic Image Noise (HU SD)

Air HU SD							
CT Scan	OHSU			Accumetra			
	Air HU SD			Air HU SD			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
1 CTLX1-B30-0.6	9.911	12.739	15.452	8.910	11.33	15.22	
2 CTLX1-B30-1.0	8.021	9.516	10.172	6.1	7.41	9.98	
3 CTLX1-B40-0.6	10.687	13.971	17.959	9.6	13.15	17.41	
4 CTLX1-B40-1.0	7.862	10.56	12.313	6.25	8.31	10.97	
5 CTLX1-B60-0.6	31.958	41.786	55.442	29.85	41.92	57.82	
6 CTLX1-B60-1.0	18.562	30.372	46.306	18.72	27.36	40.5	
7 ExtraScan	20.822	24.533	36.47	21.34	28.22	35.79	

Acrylic HU SD							
CT Scan	OHSU			Accumetra			
	Acrylic HU SD			Acrylic HU SD			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
1 CTLX1-B30-0.6	19.297	18.894	19.3	10.06	13.59	17.1	
2 CTLX1-B30-1.0	10.052	17.573	16.257	6.5	9.55	10.92	
3 CTLX1-B40-0.6	20.545	19.983	25.099	11.59	15.35	18.91	
4 CTLX1-B40-1.0	12.465	16.657	15.436	7.51	9.76	11.88	
5 CTLX1-B60-0.6	48.827	74.957	94.353	43.98	69.29	83.69	
6 CTLX1-B60-1.0	27.409	58.178	67.377	27.82	44.42	52.32	
7 ExtraScan	39.875	48.908	59.813	28.71	36.59	52.72	

2.106

-8.955

OHSU Results : Air & Acrylic Image Noise (HU SD)

B30f 0.6mm

B30f 1.0mm


B40f 0.6mm

B40f 1.0mm

B60f 0.6mm

B60f 1.0mm

LUNG 1.25mm Warped



OHSU
Accumetra

We expect differences due to high edge enhancement (noise impacted when EE is high) and warping.

OHSU Results : Spatial Warping

Z Spatial Warping							
CT Scan	OHSU Spatial Warping Y/N			Accumetra Spatial Warping Y/N			
	Module @ 0mm	Module @ 100mm	Module @ 200mm	Module @ 0mm	Module @ 100mm	Module @ 200mm	
	1 CTLX1-B30-0.6	N	N	N	0.046	0.04	0.05
2 CTLX1-B30-1.0	N	N	N	0.07	0.055	0.059	
3 CTLX1-B40-0.6	N	N	N	0.045	0.04	0.041	
4 CTLX1-B40-1.0	N	N	N	0.078	0.056	0.061	
5 CTLX1-B60-0.6	N	N	N	0.081	0.073	0.07	
6 CTLX1-B60-1.0	N	N	N	0.106	0.099	0.083	
7 ExtraScan	N	N	Y	0.12	0.284	0.432	

Any value > 0.3 is considered positive for Z spatial warping

100% agreement on spatial warping (but no visual cutoff guidance was provided)

OHSU Results : Spatial Warping

