

Quantitative
Imaging
Biomarkers
Alliance®



QIBA Profile Conformance

Self - Attestation

QIBA profile title	Diffusion – Weighted Magnetic Resonance Imaging (DWI)
QIBA profile version	December 20, 2019
Company/Institution doing self-attestation	
Company/Institution responsible person	
Date Self-Attestation was submitted to QIBA	
Date Self-Attestation was reviewed by QIBA	
Date Conformance was registered by QIBA	

Document

Some checklist items reference a required Assessment Procedure which may be found in the Profile Document.

Some checklist items have clarifications, rationale, or guidance in the corresponding Discussion section in the Profile Document.

To obtain a copy of the Profile Document, visit <http://qibawiki.rsna.org/index.php/Profiles>

If a QIBA Conformance Statement is already available for an actor (e.g. your acquisition device), a site may choose to provide a copy of that statement rather than confirming each of the requirements in that Actors checklist yourself.

Vendors publishing a QIBA Conformance Statement shall provide a set of “Model-specific Parameters” (as shown in Annex A) describing how their product was configured to achieve conformance. Vendors shall also provide access or describe the characteristics of the test set used for conformance testing.

QIBA Conformance Statements

QIBA Conformance Statements are documents prepared and published by vendors or sites to describe the intended conformance of their products, staff or institution to one or more QIBA Profiles.

Conformance requirements are defined in the QIBA Profile document for each Actor in the Profile. For some requirements, the Profile document also defines assessment procedures.

This conformance statement contains all relevant checklists for all relevant actors for site or product conformance. Supporting material is available on the QIBA wiki conformance section for the respective profile. Checklists in this conformance statement document need to be filled out.

Users can use Conformance Statements to determine whether their staff and products can be expected to deliver the biomarker performance described in the Profile Claim. Achieving the performance claim depends on all Actors described in the Profile being present at the site and conforming to the requirements.

A QIBA Conformance Statement is not intended to promote or advertise aspects of a product or site not directly related to its implementation of QIBA capabilities.

IMPORTANT NOTE: Vendors and sites are solely responsible for the accuracy and validity of their QIBA Conformance Statements. QIBA and its sponsoring organizations have not evaluated or approved any QIBA Conformance Statement or any related product, site or staff, and QIBA and its sponsoring organizations shall have no liability or responsibility to any party for any claims or damages, whether direct, indirect, incidental or consequential, including but not limited to business interruption and loss of revenue, arising from any use of, or reliance upon, any QIBA Conformance Statement.

QIBA Conformance Statement for a Product

QIBA Conformance Statement			
Vendor	Product Name	Version	Date
This product conforms to all specifications required for the QIBA Profiles and Actors listed below:			
Profiles Implemented	Actors Implemented	Notes	
Diffusion – weighted MRI 12-20-2019	Image Analysis Tools	See A.2	
Links to Additional Information			
Link to QIDW section storing phantom test data (specify data sets used for conformance testing)			
https://bit.ly/2QXLo3e			
General information on QIBA: qibawiki.rsna.org			

Annex A: Conformance Notes

A.2 DWI/ADC System QC Image Analysis Software

(for generation of performance report for MRI acquisition device assessment per 3.2.2 specifications in Diffusion – weighted MRI 12-20-2019 profile)

Parameter	Conformance (select “Yes” if any requirement is met; “No” if none are met)	Requirement
Software Deployment		
Web-based SW	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Web-based QC SW executables, updates and version numbers shall be maintained by SW vendor Optional comment _____
Web-based data	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Site QC DICOM data and derived reports shall be maintained by QC SW vendor and remain accessible to site user for specified periods from DICOM upload date: DICOM data maintained for _____ months, and QC reports maintained for _____ months following upload of DICOM data. Optional comment _____
Site-based SW installation	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	QC SW executables and updates shall be operational when installed on a site-owned server meeting the following system requirements: _____ _____ Optional comment _____
Site Data		
DWI / ADC QC phantom	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	QC SW shall be able to process data of the following phantom(s): QIBA DWI phantom at 0°C as described in https://qibawiki.rsna.org/images/a/a5/QIBA_DWI_Profile_Conformance_Testing_Supplement_1_20210401a.pdf Alternative phantom(s) _____ Optional comment _____
DWI / ADC QC scan protocol	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	QC SW shall be able to process above phantom(s) data acquired using following scan protocol(s): QIBA conformance testing as described in https://qibawiki.rsna.org/images/a/a5/QIBA_DWI_Profile_Conformance_Testing_Supplement_1_20210401a.pdf

Parameter	Conformance (select “Yes” if <i>any</i> requirement is met; “No” if <i>none</i> are met)	Requirement
		Alternative protocol(s) _____ Optional comment _____
DWI / ADC QC DICOM data requirements	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	QC SW shall be able to process above phantom(s) data acquired using above scan protocol(s) adhering to the following requirements (check all that apply): <ul style="list-style-type: none"> <input type="checkbox"/> DICOM data from GE MRI scanners <input type="checkbox"/> DICOM data from Siemens MRI scanners <input type="checkbox"/> DICOM data from Philips MRI scanners <input type="checkbox"/> DICOM data from Hitachi MRI scanners <input type="checkbox"/> DICOM data from Toshiba MRI scanners <input type="checkbox"/> DICOM data from United Imaging MRI scanners <input type="checkbox"/> Data in classic <i>DICOM format</i> <input type="checkbox"/> Data in enhanced DICOM format <input type="checkbox"/> Full exam including non-DWI series along with directional and trace DWI <input type="checkbox"/> Exam including both directional and trace DWI <input type="checkbox"/> Exam including only trace DWI <input type="checkbox"/> Exam including only scanner-generated ADC maps Optional comment _____
Site QC SW User		
Site user duties	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	For proper QC SW operation, the site user shall perform the following operations (check all that apply): <ul style="list-style-type: none"> <input type="checkbox"/> Prepare and scan phantom according to specified protocol (e.g. https://qibawiki.rsna.org/images/a/a5/QIBA_DWI_Profile_Conformance_Testing_Supplement_1_202104_01a.pdf) <input type="checkbox"/> Upload (for web-based QC SW) or navigate to (for site-based QC SW) phantom DICOM data meeting above requirements <input type="checkbox"/> Select DWI/ADC series for QC analysis <input type="checkbox"/> Manually define ROIs/VOIs on phantom target tubes for QC analysis <input type="checkbox"/> Initiate QC analysis of ROI/VOI targets <input type="checkbox"/> Generate QC analysis report(s) <input type="checkbox"/> Identify which metrics are within/outside of QIBA target performance levels

Parameter	Conformance (select “Yes” if <i>any</i> requirement is met; “No” if <i>none</i> are met)	Requirement
		<input type="checkbox"/> Identify which scan acquisition parameters are within/outside of specified scan protocol ranges <input type="checkbox"/> Initiate archival, email, or printing of QC analysis report(s) Optional comment _____

QC Analysis SW		
DICOM metadata	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	QC analysis SW shall utilize the following DICOM metadata for calculations and protocol conformance testing (check all that apply): <ul style="list-style-type: none"> <input type="checkbox"/> b-value <input type="checkbox"/> Diffusion-encode directionality and trace DWI <input type="checkbox"/> Proper pixel intensity scaling of DWI <input type="checkbox"/> Proper pixel intensity scaling and units of ADC map <input type="checkbox"/> Magnetic field strength <input type="checkbox"/> Repetition time TR <input type="checkbox"/> Echo time TE <input type="checkbox"/> Image FOVs <input type="checkbox"/> Acquisition matrices (phase and frequency) <input type="checkbox"/> Image matrices <input type="checkbox"/> Slice thickness <input type="checkbox"/> Slice center-to-center spacing <input type="checkbox"/> Quantity of slices <input type="checkbox"/> In-plane phase-encode direction (row or column) <input type="checkbox"/> In-plane parallel imaging acceleration factor <input type="checkbox"/> Number of signal averages at each b-value <input type="checkbox"/> Institution name <input type="checkbox"/> Scanner manufacturer <input type="checkbox"/> Scanner model <input type="checkbox"/> Scanner serial number <input type="checkbox"/> Scanner software version Optional comment _____
		QC analysis SW shall perform the following operations (check all that apply): <ul style="list-style-type: none"> <input type="checkbox"/> ADC maps are automatically generated on a pixel-by-pixel basis by mono-exponential fit of DWI pixel

QC SW functions	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<p>intensity vs b-value</p> <ul style="list-style-type: none"> <input type="checkbox"/> Allow definition of circular ROIs 10-20mm diameter including at least 80pixels on each sample tube <input type="checkbox"/> Location of each ROI is manually defined by the SW user <input type="checkbox"/> Location of each ROI is automatically defined by the SW <input type="checkbox"/> ROIs for each sample tube are automatically combined across slices to create a single VOI for each tube from which quantitative technical performance metrics (see below) are derived <input type="checkbox"/> Statistics are derived on a pixel-by-pixel basis over sequential multi-pass DWI series, then summarized for each ROI/VOI according to methods defined in https://qibawiki.rsna.org/images/a/a5/QIBA_DWI_Profile_Conformance_Testing_Supplement_1_20210401a.pdf <input type="checkbox"/> Location of ROIs/VOIs are rendered and numerically identified in color (in outline or pixel masks) superimposed on grayscale ADW or ADC maps for user inspection <input type="checkbox"/> ADC maps generated by SW shall provide values within random error of that calculated by the scanner or by QIBA-provided QC SW <p>Optional comment _____</p>
Technical performance metrics	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<p>Using statistical methods described in https://qibawiki.rsna.org/images/a/a5/QIBA_DWI_Profile_Conformance_Testing_Supplement_1_20210401a.pdf, the QC analysis SW shall assess MRI scanner technical performance for the following metrics (check all that apply):</p> <ul style="list-style-type: none"> <input type="checkbox"/> ADC bias at magnet isocenter <input type="checkbox"/> Within-exam ADC repeatability at isocenter (RC and wCV) <input type="checkbox"/> ADC linearity over range of known diffusion coefficient values (from 0.13 to 1.1) $\times 10^{-3}$ mm²/s or provide list of known values for the given phantom <input type="checkbox"/> ADC b-value dependence <input type="checkbox"/> ADC random error estimated from intra-exam sequential 4-pass DWI scans <input type="checkbox"/> SNR of DWI from intra-exam sequential 4-pass DWI at b=0 <input type="checkbox"/> Graphical plot and/or table of SNR of DWI estimated from intra-exam sequential 4-pass DWI at

		<p>all acquired b-values</p> <ul style="list-style-type: none"> <input type="checkbox"/> An indication of which scanner's performance metrics pass/fail thresholds established in QIBA DWI profile <input type="checkbox"/> Generated metric shall be within random error of that calculated by QIBA-provided QC SW <p>Optional comment _____</p>
QC Report output	<ul style="list-style-type: none"> <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A 	<p>The QC analysis SW shall provide exportable/printable output of the MRI scanner technical performance for the following information (check all that apply):</p> <ul style="list-style-type: none"> <input type="checkbox"/> Site name and scanner demographics including system serial number <input type="checkbox"/> Phantom scan date and DWI series numbers analyzed <input type="checkbox"/> Scan protocol compliance check-list for DICOM metadata <input type="checkbox"/> ADC bias at magnet isocenter <input type="checkbox"/> Within-exam ADC repeatability at isocenter (RC and wCV) <input type="checkbox"/> ADC linearity over range of known diffusion coefficient values (from 0.13 to 1.1) $\times 10^{-3}$ mm²/s (provide list of known diffusion values for a given phantom) <input type="checkbox"/> ADC b-value dependence <input type="checkbox"/> ADC random error estimated from intra-exam sequential 4-pass DWI scans <input type="checkbox"/> SNR of DWI from intra-exam sequential 4-pass DWI at b=0 <input type="checkbox"/> Graphical plot and/or table of SNR of DWI estimated from intra-exam sequential 4-pass DWI at all acquired b-values <input type="checkbox"/> An indication of which scanner's performance metrics pass/fail thresholds established in QIBA DWI profile <input type="checkbox"/> Location of ROIs/VOIs rendered and numerically identified in color (in outline or pixel masks) superimposed on grayscale ADW or ADC maps <p>Optional comment _____</p>