

QIBA Dynamic Contrast Enhanced (DCE) MRI Biomarker Committee (BC) Call

Monday, March 18, 2019 at 11 AM (CT)

Call Summary

Participants

Caroline Chung, MD (Co-Chair)

Hendrik Laue, PhD (Co-Chair)

Dena Flamini, RT

Wei Huang, PhD

Harrison Kim, PhD, MBA

Chen Lin, PhD

Susan Noworolski, PhD

Nancy Obuchowski, PhD

Kyunghyun Sung, PhD

RSNA

Joe Koudelik

Susan Stanfa

DCE Profile Update (Dr. Chung)

- DCE BC members were encouraged to review and comment on the updated Profile at:
<https://docs.google.com/document/d/1ytAuSsRtL2mgAIY1wq6PkVGleWvWSTdxOHGdgUu1mrA/edit?usp=sharing>
- DCE BC members were thanked for the many edits that were submitted since the March 4 call
- The goal is to finalize Section 3 and discuss Section 4 (Assessment Procedures) offline
- Dr. Kim added several comments in Section 3.13: Image Analysis
- Dr. Obuchowski reviewed and made adjustments to Section 2.1 Clinical Interpretation; text has been finalized
- Early parts of Section 3 were discussed during the March 4 call; specification tables vs. general discussion content were reviewed
- Section 3.5: Periodic QA
 - High-level T1/R1 Phantom scanning precision text remains in 3.5.1 discussion section; it describes what users need to do to be QA conformant with the Profile
 - Specifics regarding T1 or R1 reference values or ranges were moved to Section 4.1: Assessment Procedure: R1/T1 Mapping accuracy
- Section 3.6: Protocol Design
 - Signal linearity (SI) and variability between scanners to be noted in Section 4: “While published guidelines on the allowed deviation from linearity do not exist, a linear correlation coefficient between SI and R1 of 0.9 or higher is required for the Claims in this Profile”
- Dr. Chung to reach out to Dr. Shiroishi re: 3.8 clinical text/subject selection
- Discussion regarding how often a scanner should be calibrated using a phantom; recommendation to do this following a software/hardware upgrade or change in protocol
- References regarding Arterial Input Function (AIF) were added to Section 3.13: Image Analysis
- Section 3.14: Image Interpretation to be kept brief
 - Suggestion to use DRO data
 - Recommendation to identify all possible sources of measurement variability and determine the underlying Actors responsible, e.g. scanner, software, etc.
 - Have a test and set limits for amount of acceptable variability for each Actor/factor so the user will know how much precision can be expected
- Dr. Obuchowski sent signal linearity details to staff to be forwarded to Drs. Chung & Laue
- Additional text needed regarding the frequency of repeated phantom QA scanning
- Analysis section to be cleaned up regarding conformance procedures, e.g., suggested use of open-source analysis software and testing against a QIBA DRO, etc.
- Dr. Laue to send article to staff for circulation

Next DCE BC Call: Monday, April 1, 2019 at 11 AM CT

RSNA Staff attempt to identify and capture all committee members participating on WebEx calls. However, **if multiple callers join simultaneously or call in without logging on to the WebEx, identification is not possible.** Call participants are welcome to contact RSNA staff at QIBA@RSNA.org if their attendance is not reflected on the call summaries.