

QIBA Musculoskeletal (MSK) Biomarker Committee (BC) Call

Tuesday, January 26, 2021 at 10 a.m. CT

Call Summary

In attendance

Xiaojuan Li, PhD (Co-chair)

Thomas Link, MD, PhD (Co-Chair)

Michael Boss, PhD

Angie Botto-van Bemden, PhD

Robert Boutin, MD

John Carrino, MD, MPH

Majid Chalian, MD

Ruud de Boer, PhD

Maggie Fung, MEng

Ali Guermazi, MD, PhD

Peter Hardy, PhD

Jason Kim, PhD

Kecheng Liu, PhD, MBA

Annelise Malkus, PhD

Nancy Obuchowski, PhD

Yuxi Pang, PhD

Cory Wyatt, PhD

RSNA

Joe Koudelik

Susan Stanfa

Update on Special Report for *Radiology* Manuscript (Dr. Chalian)

- The manuscript has received major revisions by *Radiology* and the authors are currently working on the revisions.
- As soon as the paper is published, it will be added to the [QIBA EndNote literature collection](#)

Multisite/multivendor grant and phantom development (Dr. Li)

- Dr. Li received R01 funding to perform a multi-center standardization study, which will help advance the Profile through technical and claim confirmation
- Dr. Li has been working with Drs. Mirowski (Verellium, LLC) and Keenan (NIST) on phantom development
 - The mechanical design of a new MSK phantom has been completed
 - The team is working on finalizing the vial solution and the timeline for phantom delivery will be shared with MSK BC members

MSK Profile – Review Public Comments (Dr. Link)

- The public comment period closed on October 29, 2020, and MSK BC members have been using the [MSK public comment resolution Google Sheet](#) to document how feedback is addressed
- Forty-one of the 113 comments received have already been addressed on previous calls

Comments from Frank W. Roemer, MD (University of Erlangen-Nuremberg)

- Section 3.7: Image Data Analysis - commenter agreed that the segmentation to be overlaid to T1p and T2 maps is ideal, but in reality, this is a very challenging task that needs co-registration particularly as the 3D GRE images have much higher resolution/thinner slices compared to the T2 or T1p images
 - Suggestion to exchange “shall” for “ideally”
 - The MSK BC recommends DESS/MENSA; MAPSS should not be used for segmentation
- Section 3.7 (Figure 2) - The image marking for lateral femur will be redrawn
- Section 3.8: Image Data Interpretation (Figure 4) – an issue was noted re: normative values grouped for KLO and 1; suggestion to present and aim at displaying KLO and 1 as separate entities
 - The MSK BC clarified that the figure refers to WORMS 0 and 1, not KL 0 and 1
- Section 4.1.3: Assessment Procedure - Imaging Analysis: suggestion to explore additional data extraction beyond T2 and T1p values
 - MSK BC to consider whether to include texture analysis of cartilage T2 maps and will add references and additional discussion of problems

Comments from Gregory Chang, MD (NYU Langone and former ISMRM MSK Study Group Chair)

- Section 3.7.1: Image Data Analysis (Discussion) - suggestion to accommodate a variety of Profile user types by providing a short list of software that can perform image analysis
 - MSK BC stated that a product for cartilage segmentation is not currently available from manufacturers and will note this in this discussion section; additional information and studies will also be included
 - The Profile will be modified once manufacturers provide this as a commercial product
- Section 3.6. Image Data Acquisition Discussion – suggestion to add a recommendation for the number of echoes
 - For reproducible measurement of mono-exponential decay components, a minimum of four echoes will be recommended
 - While reproducibility for mono-exponential fitting would be good with four echoes, more echoes may provide in-depth information about cartilage degradation (e.g., providing more information about fast decaying components by using bi-exponential fitting) and introduce longer acquisition time
 - Optimized number of echoes and optimized echo spacing for specific questions are active areas of research
- Section 3.2: Installation - rheumatologists and orthopedists may be unaware that even different model 3T scanners from the same vendor could demonstrate variation in measurements, and noting this was suggested
 - MSK BC will also specify that variability occurs in the following situations as well: same vendor and field strength scanners, different models from the same vendor, and same model but different machines

Comments from Flavia M. Cicuttini, MSc, MBBS, FRACP, PhD (Monash University and OARSI board member)

- The document contains a good general approach and guidelines; however, it may need to be acknowledged that if there is the need to examine change, groups who do work on the same machine, using a standard protocol and measurement approach may be able to assess the state of the cartilage and detect clinically significant changes using modifications of the approach presented
 - MSK BC noted that the Claim is longitudinal, and the existing Profile discussion will be expanded to include modified approaches
- Section 2: Clinical Context and Claims (Important Considerations and Limitations) – The Claim requires focus on subjects with less severe cartilage loss; it was noted that a Kellgren-Lawrence (KL) score of 2 can be associated with significant (advanced) cartilage loss and evidence needs to be provided that the Claim is valid in the setting of KL2
 - OAI data are available and studies that demonstrate cartilage damage in KL2 will be cited (references need to be added in the discussion section)
- Section 2 (Discussion) - if cartilage composition changes size beyond 11-14% limits, one can be 95% confident there has been a true change in the cartilage composition
 - Clarification needed re: the difference between identifying true change in an individual compared to average change that may be detected across two groups, e.g., in a clinical trial
 - A biostatistician would be needed, and the clinical trial issue will be addressed re: detection of smaller changes; Dr. Obuchowski provided clarification

Comments from Samuel A. Einstein, PhD (York Hospital)

- Section 2: Clinical Context and Claims (Important Considerations and Limitations) - the studies supporting the Claims are based on phantoms and a few healthy volunteers (data in references are limited), yet the Claims seem to be applied to the clinic
 - Larger studies are underway, e.g., the multisite/multivendor and phantom development funded by the R01 grant; the data justifying these Claims in a clinical setting will be incorporated into the Profile

- Section 3.2: Installation - additional requirements needed for hardware specifications such as for the RF amplifier and gradient performance
 - No specific requirements are necessary because it varies from scanner to scanner and the proposed T1rho and T2 sequence lack additional requirements beyond the current standard clinical MR scanners
 - If the sequence runs on a scanner there is no need to specify RF and gradient details
- Section 3.3.1: Periodic QA (Discussion) - partial pressure of oxygen changes have a significant effect on T2 if bacteria start growing in the agarose when stored at room temperature
 - This was deemed a viable concern, which will be acknowledged in the Profile
 - Dr. Li to request input from Dr. Keenan (NIST) and Dr. Mirowski (Verellium) for a detailed response; the R01 grant proposed to check the phantom stability during the study period
 - Dr. Boss (former NIST staff) concurred that was unaware of any agarose phantoms that do not eventually degrade
 - It was also noted that NIST had very limited experience with agarose, typically using other (more stable) fill solutions with 5-year estimated stability
 - The phantoms developed during his tenure were not intended for long-term storage
- 3.3.2: Periodic QA (Specification Table) – suggestion to add ACR definition of MR scientist as a potential physicist qualification; “physicist” will appear in place of “MRI scientist” in the Stage 2: Consensus Profile draft
- Dr. Einstein’s comments will continue being addressed during the February 23 MSK BC Meeting

Next Call: Tuesday, February 23, 2021 at 10 a.m. CT [4th Tuesdays of each month]

RSNA Staff attempt to identify and capture all committee members participating on WebEx calls. However, **if multiple callers join simultaneously or call in without logging on to the WebEx, identification is not possible.** Call participants are welcome to contact RSNA staff at QIBA@RSNA.org if their attendance is not reflected on the call summaries.