

Quantitative  
Imaging  
Biomarkers  
Alliance



1  
2  
3  
4  
5  
6  
7

## Profile: DCE MRI Quantification

Version 1.0  
December 13, 2011

---

8 **Table of Contents**

9 I. Executive Summary ..... 3

10 II. Clinical Context and Claims..... 3

11     Claim: ..... 4

12 III. Profile Details..... 4

13     1. Subject Handling ..... 4

14     2. Imaging Procedure..... 8

15     3. Image Post-processing..... 12

16     4. Parametric image formation ..... 12

17     5. Parametric image analysis ..... 15

18     6. Archival and Distribution of Data ..... 18

19     7. Quality Control..... 18

20     8. Imaging-associated Risks and Risk Management..... 24

21 IV. Compliance ..... 24

22     Acquisition Scanner ..... 24

23     Contrast Inject Device..... 25

24     Software Analysis..... 25

25     Performing Site..... 25

26 References ..... 27

27 Appendices ..... 30

28     Appendix A: Acknowledgements and Attributions ..... 30

29     Appendix B: Conventions and Definitions ..... 31

30     Appendix C: Spreadsheet on reproducibility data ..... 32

31     Appendix D: Model-specific Instructions and Parameters..... 35

32

33

## 34 I. Executive Summary

35 The RSNA QIBA Dynamic Contrast Enhanced Magnetic Resonance Imaging (DCE-MRI) Technical Committee  
36 is composed of scientists representing the imaging device manufacturers, image analysis laboratories,  
37 biopharmaceutical industry, academia, government research organizations, and professional societies,  
38 among others. All work is classified as pre-competitive. The goal of the DCE-MRI committee is to define  
39 basic standards for DCE-MRI measurements and quality control that enable consistent, reliable and fit-for-  
40 purpose quantitative transfer constant ( $K^{\text{trans}}$ )<sup>[1]</sup> and blood normalized initial area under the gadolinium  
41 concentration curve ( $\text{IAUGC}_{\text{BN}}$ )<sup>[2]</sup> results across imaging platforms (at 1.5 tesla (1.5 T)), clinical sites, and  
42 time.

43  
44 This effort is motivated by the emergence of DCE-MRI as a method with potential to provide predictive,  
45 prognostic and/or pharmacodynamic response biomarkers for cancer<sup>[3-11]</sup>. Remarkably, the results  
46 demonstrating this potential have been obtained despite considerable variation in the methods used for  
47 acquisition and analysis of the DCE-MRI data. This suggests there are substantial physiological differences  
48 (i.e., benign vs. malignant or non-responsive vs. responsive tumors) underlying these observations. Thus,  
49 there appears to be a promising future for use of DCE-MRI for both clinical research and in routine clinical  
50 practice. However, in order to fulfill this promise it is essential that common quantitative endpoints are  
51 used and that results are independent of imaging platforms, clinical sites, and time.

52  
53 For the application of DCE-MRI in the development of anti-angiogenic and anti-vascular therapies, there is a  
54 consensus<sup>[12]</sup> on which quantitative endpoints should be employed:  $K^{\text{trans}}$  and  $\text{IAUGC}_{\text{BN}}$ . Hence, the initial  
55 focus of the DCE-MRI committee is on these biomarkers. Although there have been general  
56 recommendations on how to standardize DCE-MRI methodology<sup>[12, 13]</sup>, there are no guidelines sufficient to  
57 ensure consistent, reliable and fit-for-purpose quantitative DCE-MRI results across imaging platforms,  
58 clinical sites, and time. Hence, in this profile, basic standards for site and scanner qualification, subject  
59 preparation, contrast agent administration, imaging procedure, image post-processing, image analysis,  
60 image interpretation, data archival and quality control are defined to provide that guidance.

### 61 ***Summary of Clinical Trial Usage***

62 *This technique offers a robust, reproducible measure of microvascular parameters associated with human*  
63 *cancers based on kinetic modeling of dynamic MRI data sets. The rigor and details surrounding these data*  
64 *are described throughout the text of this document in various sub-sections.*

## 65 II. Clinical Context and Claims

66 One application of DCE-MRI where considerable effort has been focused on quantitative endpoints is its use  
67 to provide pharmacodynamic biomarkers for the development of novel therapeutic (in specific anti-  
68 angiogenic) agents targeting the tumor blood supply<sup>[4, 9, 14-26]</sup>. A growing understanding of the underlying  
69 molecular pathways active in cancer has led to the development of novel therapies targeting VEGFR, EGFR-  
70 tk, PI3K, mTOR, Akt and other pathways. Unlike the conventional cytotoxic chemotherapeutic agents, many  
71 of these molecularly-targeted agents are cytostatic, causing inhibition of tumor growth rather than tumor  
72 regression. One example is anti-angiogenesis agents, which are presumed to act through altering tumor  
73 vasculature and reducing tumor blood flow and/or permeability. In this context, conventional endpoints,  
74 like tumor shrinkage as applied at e.g. Response Evaluation Criteria in Solid Tumors (RECIST), may not be

the most effective means to measure therapeutic responses. Other functional MR imaging acquisition and analysis applications (e.g. BOLD,  $R_2^*$  perfusion) yield several important candidate imaging biomarkers that can predict and monitor targeted treatment response and can document pharmacodynamic response. However, these are not within the scope of this document. DCE-MRI represents an MRI-based method to assess the tumor microvascular environment by tracking the kinetics of a low-molecular weight contrast agent intravenously administered to patients.

The emerging importance of angiogenesis as a cancer therapy target makes assays of vascularity important to clinical research and future clinical practice related to targeted cancer therapy. There are multiple literature reports of the application of DCE-MRI to predict and detect changes associated with angiogenesis targeted therapy<sup>[4, 9, 15, 17, 19, 20, 24, 25]</sup>. Further, there is interest in the application of quantitative DCE-MRI to characterize enhancing lesions as malignant in several organ systems, including breast and prostate.

In this context,  $K^{\text{trans}}$  and IAUGC<sub>BN</sub> can provide evidence of the desired physiologic impact of these agents in Phase 1 clinical trials. For some agents, e.g., VEGFR-targeted agents, evidence of substantially reduced  $K^{\text{trans}}$  and IAUGC<sub>BN</sub> is necessary, but not sufficient, for a significant reduction in tumor size<sup>[16, 17]</sup>. For other agents, e.g., vascular-targeted agents, evidence of a substantial vascular effect may not be associated with a reduction in tumor size<sup>[9]</sup>, but is still essential for effective combination with other anti-cancer agents. In either case, lack of a substantial vascular effect indicates a more potent agent is needed, while evidence for a substantial vascular effect indicates further development is appropriate.

### **Utilities and Endpoints for Clinical Trials**

*DCE-MRI is currently not the standard of care in many centers conducting clinical trials in oncology. Since these centers often do not have expertise in DCE-MRI and more than one center is typically involved, therefore effort and precision are required to ensure consistent, reliable and fit-for-purpose quantitative DCE-MRI results. Hence, the guidelines provided in this profile will ensure that not only are the relative changes induced by treatment are informative, but that absolute changes can be compared across these studies.*

#### **Claim:**

**Quantitative microvascular properties, specifically transfer constant ( $K^{\text{trans}}$ ) and blood normalized initial area under the gadolinium concentration curve (IAUGC<sub>BN</sub>), can be measured from DCE-MRI data obtained at 1.5T using low molecular weight extracellular gadolinium-based contrast agents with a 20% within-subject coefficient of variation for solid tumors at least 2 cm in diameter.\***

Profile specified for use with: **patients with malignancy**, for the following indicated biology: **primary or metastatic**, and to serve the following purpose: **therapeutic response**.

\* a 20% within-subject coefficient of variation is based on a conservative estimate from the peer-reviewed literature. In general, this suggests that a change of approximately 40% is required in a single subject to be considered significant.

## **III. Profile Details**

### **1. Subject Handling**

---

## 1.1 Subject Scheduling

### *Subject Selection Criteria related to Imaging*

- Local policies for contraindications for absolute MRI safety should be followed; definition of relative and/or absolute contraindications to MRI are not within the scope of this document.
- Lesions that are selected for DCE-MRI analysis should not be within 10 cm of metal prostheses, e.g., spinal hardware, hip prostheses, metallic surgical staples, etc.
- Patient selection criteria may be guided by the Eastern Cooperative Oncology Group (ECOG) status (See Appendix B) for full description of ECOG performance status). In specific, patients meeting ECOG status  $\geq 2$  will not be eligible for participation in the study because, historically, this patient profile has shown poor ability to meet the demands of the examination.
- The QIBA DCE-MRI committee acknowledges that there are potential and relative contraindications to MRI in patients suffering from claustrophobia. Methods for minimizing anxiety and/or discomfort are at the discretion of the physician caring for the patient.
- The QIBA DCE-MRI committee acknowledges that there are potential risks associated with the use of gadolinium-based contrast media. The default recommendations for intravenous contrast that follow assume there are no known contraindications in a particular patient other than the possibility of an allergic reaction to the gadolinium contrast agent. The committee assumes that local standards for good clinical practices (GCP) will be substituted for the default in cases where there are known risks.
- Recent FDA guidelines (<http://www.fda.gov/Drugs/DrugSafety/ucm223966.htm#approved>), outline the safety concerns associated with using gadolinium based contrast agents in patients with impaired renal function. The DCE-MRI committee echoes these recommendations and advises reference to these standards when choosing patients in order to determine eligibility for entry into a DCE-MRI clinical trial.
- Although the vascular half-life of the gadolinium contrast agents addressed by the Profile is approximately 90 min, it is strongly recommended that patients should not have received ANY gadolinium based contrast agent within 24 hrs before a DCE-MRI procedure as some residual contrast agent may remain in the lesion(s) of interest and the impact of such residual contrast agent on the within-patient coefficient of variation is unknown.

#### **1.1.1. Timing of Imaging Tests within the Treatment Calendar**

The DCE-MRI Technical Committee believes that all baseline evaluations should ideally be within 14 days. Otherwise the resulting functional tumor characterization may not reflect the status of the tumor prior to initiation of therapy. The interval between follow up scans within patients may be determined by current standards for GCP or the rationale driving a clinical trial of a new treatment

---

**1.1.2. Timing Relative to confounding Activities (to minimize “impact”)**

The presence of susceptibility artifacts and, possibly, mass-effect from hemorrhage and/or air related to recent biopsy may potentially affect the quantitative DCE-MRI parameters. If practical, it is recommended that DCE-MRI examinations should not be performed within 14 days after biopsy of lesions of interest. If this amount of delay is impractical, excluding hemorrhagic portions of lesions from the image analysis is strongly recommended.

**1.2. Subject Preparation**

There are no specific patient preparation procedures for the MRI scans described in this protocol. There are specifications for other procedures that might be acquired contemporaneously, such as requirements for fasting prior to FDG PET scans or the administration of oral contrast for abdominal CT. Those timing procedures may be followed as indicated without adverse impact on these guidelines

**1.2.1. Prior to Arrival**

The local standard of care for acquiring MRI scans may be followed. For example, patients may be advised to wear comfortable clothing, leave jewelry at home, etc.

**1.2.2. Upon Arrival**

Staff shall prepare the patient according to the local standard of care, (including e.g. removal of all metal objects and electronic devices). Patients should be comfortably positioned, in appropriate attire to minimize patient motion and stress (which might affect the imaging results) and any unnecessary patient discomfort.

**1.2.3 Preparation for Exam**

Beyond a clear, simple language description of the image acquisition procedure, no exam preparation is specified beyond the local standard of care for MRI with contrast.

**1.3. Imaging-related Substance Preparation and Administration****1.3.1. Substance Description and Purpose**

The literature, which supports the claim, is based on the utilization of an extracellular gadolinium based contrast agent. Although it is known that there is a small degree of protein binding associated with many commercially available extracellular gadolinium contrast agents,<sup>[27]</sup> these are comparable amongst the various vendors. Contrast agents with fundamentally different degrees of protein binding, (e.g., Gadobenate and Gadofosveset) are not addressed by this profile. The committee therefore recommends using a classical extracellular based gadolinium based contrast agent.

**1.3.2. Dose Calculation and/or Schedule**

Total contrast agent dose depending on body weight and renal function:

- 
- 189 • Before DCE-MRI the patient's renal creatinine clearance should be obtained, and estimated  
190 glomerular filtration rate (eGFR) determined through well-known and adopted formulas.<sup>[28]</sup>  
191 • Routine dose of the Gadolinium contrast agent should be 0.1 mmol/kg.  
192 • The decision whether to administer total contrast dosage will be based on GCP and the  
193 policies adopted at the institution performing the examination. However, the same body weight adapted  
194 contrast agent concentration should be used for repeat studies, and in case of an acute renal insufficiency  
195 and/or failure at follow-up a later imaging time point or patient exclusion should be discussed.

### 196 **1.3.3. Timing, Subject Activity Level, and Factors Relevant to Initiation of Image Data Acquisition**

197 Contrast injection should occur after the following imaging sequences have been acquired (See Section 6):

- 198 • Anatomic imaging for localizing tumors  
199 • Variable flip angle imaging for native tissue (pre-gadolinium injection) T<sub>1</sub> map calculation

200 Contrast injection should occur after at least 5 baseline acquisitions from the imaging volume have been  
201 acquired.

### 202 **1.3.4. Administration Route**

203 It has been demonstrated in studies of CT arteriography, contrast-enhanced CT, and contrast-enhanced MR  
204 arteriography that left arm injections lead to reflux of contrast agent into venous structures<sup>[29-31]</sup> It stands  
205 to reason that inconsistencies in the arm that is injected could, therefore, lead to variability in the shape of  
206 the VIF, further exaggerating the potential inaccuracy of an assumed input function. Therefore, it is  
207 recommended that each subject should have an intravenous catheter (ideally no smaller than 20 gauge  
208 (0.8mm inner diameter)), which should be ideally placed in the right antecubital fossa. Injection through a  
209 port-a-catheter or permanent indwelling catheter is not recommended. What is critical is that the same  
210 injection site and catheter size be used for repeat studies, if at all possible.

### 211 **1.3.5. Rate, Delay and Related Parameters / Apparatus**

212 Contrast agent and normal saline flush should be administered in a dynamic fashion with an MR-compatible  
213 power injector.

- 214 • At baseline and at each subsequent time-point in any longitudinal study, the same dose of contrast (in  
215 mmol/kg) and rate of contrast administration should be performed.  
216 • The rate of administration should be rapid enough to ensure adequate first-pass bolus arterial  
217 concentration of the contrast agent (generally 2-4 ml/sec)  
218 • The contrast agent should be flushed with 20 to 30 ml of normal saline, which should be injected at the  
219 same rate as the contrast agent.

### 220 **1.3.6. Required Visualization / Monitoring, if any**

221 No particular visualization or monitoring is specified beyond the local standard of care for MRI with  
222 contrast.  
223  
224  
225

## 2. Imaging Procedure

This section describes the imaging protocols and procedure for conducting a DCE-MRI exam. Suitable localizer (scout) images must be collected at the start of exam and used to confirm correct coil placement as well as selection of appropriate region to image. This will be followed by routine non-contrast agent-enhanced sequences to delineate the number, location, and limits of tumor extension. Exact protocols for these imaging sequences may be determined by the local imaging norms, e.g:

- **Localizer**
- **Anatomic sequences  $T_1$ ,  $T_2$  weighted imaging**
- **Variable Flip angle (VFA)  $T_1$  weighted imaging ( $T_1$  mapping)**
- **3D Gradient echo volumetric imaging (dynamic imaging)**
- **Anatomic, post-contrast  $T_1$  weighted sequences**

### 2.1. Required Characteristics of Resulting Data

The DCE-MRI portion of the exam will consist of two components, both acquired using the same 3D fast spoiled gradient recalled echo sequence, or equivalent, and scan locations:

(a) A variable flip angle (VFA) series, for pre-contrast agent native tissue  $T_1$  mapping.

- Ensure TR and TE values stay constant for all flip angles,
- Ensure that the machine gain settings are not reset automatically (using automated pre-scan features) between each flip angle acquisition so that system gain settings are identical for each flip angle acquisition.
- Flip angles: The range of numbers of flip angles supported in the literature varies from 2-7.
- Number of signal averages (NSA or NEX)  $\geq 2$ .
- Fat saturation if used may alter baseline  $T_1$  values and therefore should be consistently used throughout the examination.
- The pulse sequence and coils used for  $T_1$  calculation should be the same used for the DCE-MRI Protocol (see 2.1 b).

(b). DCE-MRI Protocol: Pulse Sequence:

- Pulse Sequence:** 3D fast spoiled gradient recalled echo or equivalent
- Coils:** Transmit: Body coil; Receive: Body coil or phased array receive coil dependent on which body part is being studied, e.g., torso (pelvic applications), breast coil (breast applications)
- Parallel imaging options are not recommended due to vendor-specific implementations of such techniques and the fact that the effects of such techniques on within-patient coefficients of variation in  $K_{trans}$  and  $IAUGC_{BN}$  have not been evaluated.
- No magnetization preparation schemes are specifically addressed by this Profile, including the use of saturation pulses for fat suppression. The use of such pulses may impact the within-subject coefficient of variation and should be investigated prior to use.
- **Imaging plane** - The acquisition plane should include the lesion of interest and a **feeding vessel with in-plane flow** in order to capture a **vascular input function (VIF)**. In addition, the choice of the acquisition plane should be made, where possible, to mitigate the effects of



lesion motion, e.g., coronal-oblique plane for a liver lesion.

- **Frequency encoding direction:** The frequency encoding direction should be adjusted so as to minimize motion artifact. This decision will be based on the location of the tumor being interrogated and its relationship to moving structures.

Parameter	Compliance Levels (for DCE acquisitions)	
TE	Acceptable	2.0-2.5ms
	Target	1.5-2.0ms
	Ideal	<1.5ms
TR	Acceptable	5-7ms
	Target	3-5ms
	Ideal	< 3ms

\*Note: The table above specifically addresses the DCE-MRI acquisition. The choices of TE and TR might be modified slightly for the pre-gadolinium administration R1 measurements. For example, the TR may be lengthened for more optimal R1 quantification.

- **Temporal resolution:** The temporal resolution should be less than 10 sec.
- **Flip angles:** Flip angles ranging from 25-35 degrees are recommended in order to minimize saturation effects. Smaller flip angles will lead to potential saturation of the signal intensity vs. gadolinium concentration, particularly in vessels. It should be noted that SAR limits may affect the maximum allowable flip angle and, of course, such limits may be affected by the patient’s weight and, for some scanners, weight and height. The technologist should use the maximal allowed flip angle when SAR limitations occur. In addition, the number of imaging sections may be reduced, if practical, to help mitigate the SAR limitations while maintaining a flip angle in the desired range stated above.
- **Receiver Bandwidth:** Greater or equal to ±31.25 kHz (or ~250 Hz/pixel)
- **Field of View (FOV) and Partial Fourier (“fractional echo” and/or reduced phase-encoding FOV) as needed to meet temporal resolution requirements**
- **Number of Slices:** Acceptable: ≥10 prior to zero fill. Ideal: as many as possible while maintaining ideal temporal resolution.
- **Slice thickness:** *Ideal:* <5 mm, *Target:* 5.1-6 mm, *Acceptable:* 6.1-8 mm
- **Matrix:** 256 x 160 (before applying rectangular FOV) – in order to meet 1-2mm in-plane

---

298 spatial resolution

- 299
- 300 • **Number of acquisitions (phases):** Sufficient to allow acquisition of at least 5 minutes of post  
301 injection data plus at least 5 phases acquired before contrast agent injection (baseline  
302 images).
  - 303
  - 304 • **Digitized bit depth:** The maximum dynamic range should be utilized, e.g., “extended  
305 dynamic range” or equivalent.

### 306 **2.1.1. Data Content**

307 All imaging data should be stored in DICOM format.

### 308 **2.1.2. Data Structure**

309 All variable flip angle (VFA) data should be clearly labeled as individual series, one per flip angle, or  
310 contained in a single series with the data order clearly defined.

311

312 All DCE-MRI data should be contained in a single series.

### 313 **2.1.3. Data Quality**

314 A quality review, confirming that all imaging parameters were correct, data structure is correct, etc., before  
315 the data are submitted for analysis.

## 316 **2.2. Imaging Data Acquisition**

### 317 **2.2.1. Subject Positioning**

#### 318 **(a) Patient and coil positioning:**

- 319
- 320 • When the general location of the target tumor(s) is known prior to DCE-MRI, for example glioma or  
321 local breast cancer evaluation, the patient set up for the MRI should be based on standard operating  
322 procedures for patient positioning and coil placement for clinical MRI of that body part taking into account  
323 the total scan time (see below).
  - 324
  - 325 • When the subject under investigation may have uncertain tumor location(s), as is common in the  
326 setting of patients undergoing therapy for metastatic disease, it will often be necessary for the DCE-MRI  
327 study to be planned with reference to the most recent pre-DCE-MRI imaging (often a CT study). From this  
328 study, tumor burden and location should be assessed. Optimally, review of actual imaging by a radiologist  
329 involved in the DCE-MRI study planning should be made. At times, if such images are not available for direct  
330 review, review of imaging reports (CT, PET) detailing extent of disease is mandatory, both to confirm  
331 eligibility (presence of at least one “imageable” target lesion) and to identify the preferred anatomic  
332 regions for DCE-MRI (chest, abdomen, pelvis, extremity). Review of prior diagnostic imaging may also be  
333 helpful to confirm cystic or necrotic nature of certain lesions, assessments which may be challenging at the  
334 time of DCE-MRI planning based solely on T<sub>1</sub>- and/or T<sub>2</sub>-weighted image sets. When multiple potential

target lesions are available, the location of the most suitable lesion(s) should be noted. The most suitable lesion will depend on size, location relative to areas of pulsatile or respiratory artifacts, and presence or absence of necrosis or cystic areas.

- DCE-MRI subject should be placed appropriately in the scanner in order to best image the lesion of interest (e.g. supine for head/neck/thorax/abdomen/pelvis and prone within a breast coil for breast studies).

- When patient condition allows, placement of the arms over the head may avoid undesirable wrap artifact for temporally optimized 3D spoiled gradient echo sequences used for chest and abdomen lesions. However, these positions often cannot be sustained by patients without excessive discomfort. In such cases, arms placed anteriorly over the chest or at the sides may be preferable. For larger patients, side-down arm positioning may require adjustment of the DCE-MRI imaging FOV to avoid undesirable wrap artifact. Appropriate coil placement per area of examination (head, neck, breast, extremity) is then done. For lesions in the chest, abdomen, or pelvis, a torso array coil is then placed in the area of target lesion(s). Ideally, both anterior and posterior coils are centered over the expected target lesion location.

- Tumor size and location on longitudinal studies should be considered in the design of an analysis scheme. Recall, that the claims of this profile are only applicable to lesions greater than or equal to 2cm. If the lesion is large in proportion to the volume imaged by DCE-MRI, precautions should be taken to maximize the possibility that the same portion of the lesion will be imaged on longitudinal studies. In general, this requires careful scan location set up on follow-up studies in order to match the same anatomic positions imaged in target organs on earlier studies (e.g. by saving of the planning screen shot). However, because of differences in patient angulation on follow-up studies the same anatomic locations may not be imaged on each study. In this case, an analysis scheme that discards image data from locations that are not included in the imaged volume (after end slice elimination) of all relevant studies is favored. This can be accomplished by registration of images obtained from the dynamic sequences of all studies (for example, images obtained by averaging all dynamic images obtained at the same location) to high-resolution anatomic images obtained (for example) at the initial time point.

- Tumors that are predominantly solid without significant necrosis or cystic characteristics would be considered the ideal choice of tumor for analysis. Tumors with extensive hemorrhage, or completely cystic or necrotic lesions are considered non-ideal and should be excluded from consideration.

- Tumor locations should be chosen to minimize the effects of excessive respiratory or pulsatile motion. Ideally, these would include the soft tissues of the extremities, posterior chest wall, retroperitoneum and abdomen. Although areas with some respiratory motion (e.g. kidneys, adrenal glands, retroperitoneum, lateral chest wall, pancreas, lung apices, neck) are considered acceptable, lesions within the hila, pericardium and lateral segment of the left lobe of the liver are not ideal because of their significant compromise secondary to respiratory motion.

### **2.2.2. Instructions to Subject During Acquisition**

The patient will be instructed to relax and perform slow, steady breathing during the examination.

### **2.2.3. Timing/Triggers**

378 All examinations will be performed in slow free breathing state. Timing parameters for the bolus injection  
379 of contrast agent will occur after the acquisition of no less than 5 baseline volume scans.

#### 380 **2.2.4. Model-specific Parameters**

381 Appendix D lists acquisition parameter values for specific models/versions that can be expected to produce  
382 data meeting the requirements of Section 7.1.

### 383 **2.3. Imaging Data Reconstruction**

384 All imaging data reconstruction will be performed per vendor specification and will involve Fourier  
385 transformation of Cartesian data. No user-selected smoothing or other post-processing will be performed  
386 so as to insure the integrity of the data for image analysis.

#### 387 **2.3.1. Platform-specific Instructions**

388 Appendix D lists reconstruction parameter values for specific models/versions that can be expected to  
389 produce data meeting the requirements of Section 7.2.

## 390 **3. Image Post-processing**

391 There are no specific image post-processing requirements in this profile. No user-selected post-processing  
392 filters or image normalization methods should be used prior to data analysis as described in the next steps.  
393 If phased-array receiver coils are used, image combination and reconstruction should be according to  
394 standard manufacturer algorithms.

## 395 **4. Parametric image formation**

396 Analysis of DCE-MRI data is carried out in a series of distinct steps:

397

- 398 • Generate a native tissue  $T_1$  map using the VFA data.
- 399 • When required, apply time-series motion correction to the dynamic data.
- 400 • Convert DCE-MRI signal intensity data,  $SI(t)$ , to gadolinium concentration ( $[Gd](t)$ ).
- 401 • Calculate a vascular input function.
- 402 • Identify the region or regions of interest in the dynamic data.
- 403 • Calculate the DCE-MRI imaging biomarker parameters,  $K^{trans}$  and  $IAUGC_{BN}$ .

404 Each of these steps is addressed in detail below.

### 405 **4.1. Input Data to Be Used**

406 Processed magnitude images will be utilized for image analysis for input into the steps described in the  
407 following sections

### 408 **4.2. Methods to Be Used**

#### 409 **(a) Generate a $T_1$ Map**

410 The intent of this step is to provide a complete map of pre-contrast  $T_1$  values for the imaged slab. These  
411 values will then be used in the signal formation model based conversion of changes in signal intensity to

gadolinium concentration. The slice locations, orientation and resolution of these images should be prescribed identically to the dynamic series, and this series should be acquired immediately prior to the dynamic series. The output of this step is an image of  $T_1$  values which can be co-registered to the dynamic series and used in subsequent calculations. The  $T_1$  values at each voxel location are calculated as follows [1]:

1. Create a vector  $x$  containing the signal intensity at each flip angle divided by the tangent of the flip angle.
2. Create a vector  $y$  containing the signal intensity at each flip angle divided by the sine of the flip angle.
3. For the  $n$  acquired flip angles create a set of points  $(x_0, y_0) \dots (x_n, y_n)$ .
4. Fit a line with slope  $s$  to the set of points defined in Step 3.
5.  $T_1 = -TR/\ln(s)$ .

The use of non-linear curve fitting methods (for example, simplex or Levenberg-Marquard techniques) to extract  $T_1$  from the signal intensities theoretically may be more robust to noise than the linearized solution presented above. Non-linear techniques may be used if they are validated using test images to perform no worse than the solution above in the expected range of  $T_1$ , equilibrium magnetization and noise of tumors and vessels to be imaged.

#### **(b) Apply Motion Correction to the Dynamic Data**

The intent of this step is to correct for patient motion that occurs between acquired phases of the dynamic data due to respiration, swallowing, and other involuntary movements. This step is not intended to correct ghosting artifacts that can occur along the phase encoding direction within a particular image due to patient motion during acquisition. These artifacts are more or less intractable unless the motion is regular and easily modeled, and are best addressed by adjusting the phase/frequency encoding scheme to minimize their impact on structures of interest. In general, simple rigid shift or affine transform based registration methods will not be adequate for this step, due to the fact that the movement in question is typically limited to specific regions within the image – for example, the liver in a coronal scan of the abdomen may move substantially with respiration while the bulk of the body remains relatively motionless. Fully deformable registration methods based on optical flow may provide good results in some cases<sup>[32, 33]</sup>. However, these methods will frequently fail for the phases immediately surrounding the contrast injection. Semi-automated registration in which a user identifies the target tumor and only information drawn from that region is used to generate phase to phase shifts provides an alternative approach. This allows rigid shift methods using mutual information<sup>[34]</sup>, which tend to be more robust than optical flow methods, to be employed. Finally, registration may be carried out manually or using simple shift registration techniques<sup>[21]</sup>. Data corrupted with motion must be either corrected prior to analysis or discarded for subsequent pharmacokinetic analysis.

#### **(c.) Convert $SI(t)$ in the Dynamic Data to $[Gd](t)$**

The intent of this step is to convert the arbitrary signal intensity units in the dynamic data into units of gadolinium concentration. This step should be applied after the regions of interest for analysis have been defined, but prior to the calculation of vascular parameters. Two methods for accomplishing this are defined below.

**Method A: Conversion Using a Signal Formation Model Gadolinium concentration at each image pixel is**

458 **given by (eq 1):**

$$459 \quad C(t) = \left( \frac{1}{T_1(t)} - \frac{1}{T_{10}} \right) / R_{Gd} \quad \text{Eq. 1}$$

460 Here  $T_{10}$  is the pre-contrast  $T_1$  at that pixel, obtained as described above, and  $R_{Gd}$  is the relaxivity of Gd  
461 (obtained from contrast agent manufacturer's specifications).

462  $T_1(t)$  can be derived from the SPGR imaging equation (neglecting  $T_2^*$  effects, assuming  $T_2^* \gg TE$ ) and is  
463 given by the following expressions (eqs 2-4): Let

$$464 \quad E_{10} = \exp(-TR/T_{10}) \quad \text{Eq. 2}$$

$$465 \quad B = \frac{1 - E_{10}}{1 - \cos \alpha * E_{10}} \quad \text{Eq. 3}$$

$$466 \quad A = B * SI(t) / SI(0) \quad \text{Eq. 4}$$

467 where  $\alpha$  is the flip angle, TR is the repetition time, and SI(t) and SI(0) are the signal intensities at time t and  
468 pre-contrast baseline respectively in the DCE-MRI sequence (eq 5). Then,

$$470 \quad \frac{1}{T_1(t)} = \frac{-1}{TR} * \ln \left[ \frac{1 - A}{1 - \cos \alpha * A} \right] \quad \text{Eq. 5}$$

#### 472 **Method B: Conversion Using a Look-Up Table**

473  
474 This method is motivated by the concern that inaccuracies in  $T_1$  mapping and/or co-registration of initial  $T_1$   
475 values to the dynamic data may introduce excessive variability into the final calculated parameters. If this  
476 method is used, it is not necessary to acquire the  $T_1$  mapping data described above. This method assumes a  
477 high degree of response uniformity, and so may be limited in cases where phased array coils are used. In  
478 general it is recommended to use the inherent body coil for both transmit and receive when using this  
479 method. It should also be noted that the use of this method will introduce a uniform bias in the estimation  
480 of quantitative parameters which will impact absolute measurements, but will not affect quantification of  
481 change, for example from one exam to another. This method has been shown to yield better test-retest  
482 reproducibility than  $T_1$ -based quantification method. <sup>[14, 35]</sup>

483  
484 This method requires that a phantom containing a range of concentrations of gadolinium and a range of  
485 baseline  $T_1$  values (generally obtained via different concentrations of copper sulfate or a similar compound)  
486 is scanned using the dynamic protocol on each scanner that will be used for the study. Data from these  
487 phantoms can then be used to construct a look-up table relating baseline  $T_1$ , signal delta, and gadolinium  
488 concentration. In order to create this look-up table, a linear correlation is performed between the  
489 difference of signal intensity between that in a phantom concentration sample and a sample with no  
490 gadolinium concentration (used as x-axis values) and the nominal  $R_1$  ( $1/T_1$ ) of the concentration sample.  
491 The resulting slope  $m$  then be used to estimate Gd concentration  $C$  using the equation  $C = m * [SI(t) - SI(0)]$ ,  
492 where SI(t) is the signal intensity in the dynamic data for a given time point t, and SI(0) is the signal intensity  
493 in the same location at baseline (before contrast agent injection).

#### 495 **(d) Calculate a Vascular input Function (VIF)**

The intent of this step is to generate an accurate, patient-specific vascular input function (VIF) to serve as an input to the vascular model. One way to accomplish this is to have an analyst draw a manual ROI within an artery, and use the mean enhancement curve within that ROI as the subject-specific VIF, as described by Vonken et al. [36]. It has been demonstrated previously that this method has significant variability associated with it [37], due primarily to the spatially- and temporally-varying flow artifacts found in major arteries. A better option is to make use of an automated search technique to generate a locally optimal VIF. Several methods of accomplishing this have been described previously [38-40]

The signal for the vascular input function can then be converted into concentration using either Method A or B as described above.

In some cases, data driven vascular input functions may be difficult to measure accurately due to anatomy, motion, flow effects, and  $T_2^*$  effects. In these situations, alternative methods of using population averaged vascular input functions [41-44] or reference tissue based vascular input functions [41-44] may be used. These methods in general lead to poorer characterization of subject-specific physiology and lead to poorer reproducibility [45].

#### (e) Calculate the Vascular Parameters

The intent of this step is to generate the parameter set which will be used to characterize the tissues of interest. Parameters will be calculated based on the standard Tofts model [39], which is derived from the Kety equations [46]. The vascular bed is modeled as a linear system, such that (eq 6):

$$C_t(t) = C_p(t) \otimes h(t) \quad \text{Eq. 6}$$

with impulse response  $h(t)$  given by (eq 7):

$$h(t) = K^{trans} * \exp(-k_{ep}t) \quad \text{Eq. 7}$$

where  $K^{trans}$  is the volume rate constant between blood plasma and extra-cellular extra-vascular space (EES) and  $k_{ep}$  is the rate constant between the EES and blood plasma. Given the tissue uptake curve  $C_t(t)$  and the VIF  $C_p(t)$ ,  $K^{trans}$  and  $k_{ep}$  are estimated using a gradient-descent energy minimization scheme, by using already established Levenberg-Marquardt or Minpack-1 curve fitting algorithms, both of which require adequate baseline sampling [47]. Delay correction should be performed to shift the VIF curve to match the arrival time of the tumor curve for each voxel prior to curve fitting.

A full parameter set will be calculated for each voxel within the defined tumor boundaries. Parameters may be reported out either as mean and median values per tumor or as histograms.

The baseline timepoint is defined as the timepoint immediately preceding the change in gadolinium concentration intensity. The blood normalized IAUGC<sub>BN</sub> is defined as the area under the concentration curve from the baseline timepoint up to 90 seconds post bolus arrival within the tumor, divided by the area under the vascular input function curve, up to 90 seconds post the baseline timepoint within the vessel.

#### 4.4. Platform-specific Instructions

Appendix D lists image analysis parameter values for specific models/versions that can be expected to produce data meeting the requirements of Section 5.

### 5. Parametric image analysis

Derivation of quantitative parameters characterizing the response associated with a lesion of interest from

parameter maps is a multistep process, most, if not all, of which are being studied by on-going research. There are several choices that can be made at any of these steps, and the effect of these choices on the validity of results and variability of parametric maps has not yet been fully characterized.

When multi-institutional trials are undertaken, a central site for analysis is highly recommended so as to reduce variability in analysis.

### 5.1. Input Data to Be Used

The input data that will be utilized will be in the form of concentration curves, and parametric maps of  $K^{\text{trans}}$  and  $\text{IAUGC}_{\text{BN}}$  from which ROI analysis can be performed. One shortcoming of the 3D fast spoiled gradient recalled echo technique used to acquire the dynamic images is that initial and end slice locations give inaccurate results due to wraparound artifact and variability in excitation profile. The extent of this wraparound artifact is dependent on slice-oversampling and other vendor specific techniques. Image analysis can begin by removing areas that are subjectively compromised by wraparound artifact. One method that can be used to determine which slices to discard is to closely examine the  $T_1$  maps obtained at the initial and end slice locations. Marked non-physiologic overestimations of  $T_1$  on initial and end slices are indicative of artifact.

### 5.2. Methods to Be Used

The following methodology for image interpretation of parametric maps should be performed in order to ensure complete reproducible and interpretable results.

#### (a) *Tumor ROI Definition.*

- The first step in the extraction of quantitative parameters ( $K^{\text{trans}}$  or  $\text{IAUGC}_{\text{BN}}$ ) associated with a particular lesion is to segment this lesion from adjacent tissues. Which techniques of segmentation are ideal or even acceptable for a given application is the subject of on-going research, but it is clear that the segmentation techniques used must be tailored to the particular organ system being studied with DCE-MRI. The following guidelines are proposed:
- The committee does not recommend an analysis scheme where an operator defines a lesion by placing regions of interest directly on parameter maps as that will introduce bias into the results
- Less subjective results can be obtained by using correlative imaging to define the lesion. These correlative images may be obtained at the same imaging session but not directly related to the DCE-MRI images. (For example, a  $T_2$ -weighted image of an organ, which clearly delineates lesions and their boundaries, may be used.) Correlative images should be obtained in the same imaging plane as the DCE-MRI series, with similar FOV and spatial resolution, if feasible. In this scenario, a registration step will likely be required (see 9.2)
- An alternative approach, which may be helpful if the lesion is well delineated on contrast-enhanced  $T_1$ -weighted MRI, is to create summation images (images obtained by adding together images obtained on the dynamic series for each slice location). The average images can be used to segment the lesion on one or more slices, and because these segmentations



581 are (in the absence of patient or organ motion) registered to the dynamic series, the  
582 segmentations can be used to directly extract lesion-based parameters from parametric  
583 maps.

- 584
- 585 • Because of the presence of image noise on source images of the dynamic series, along with  
586 time-dependent changes in signal intensity which may blur or even obliterate the border  
587 between lesion and background tissue, analysis schemes in which lesions are segmented  
588 independently on each image of the dynamic series should be avoided where possible. In the  
589 case of moving organs, it may be necessary to segment the lesion of interest on early  
590 (preferably, before the arrival of the contrast bolus) or late dynamic images and estimate the  
591 position of the segmented lesion in intermediate time points.
- 592
- 593 • Although lesions can be segmented using manual techniques, several techniques are  
594 available that allow a semi-automated approach to be used. The training of operator or  
595 operators in performing segmentations should be documented, preferably with training sets.
- 596

597 **(b) Registration of segmentations and parameter maps.**

598

599 Unless the segmentations are derived from relatively motion-free or motion-corrected dynamic images (for  
600 example, summary images) image registration techniques may need to be used to place the segmentations  
601 and parameter maps into a single anatomic framework (see Section 4.2). The choice of registration  
602 technique to be used depends upon the organ system being imaged; the details of this are beyond the  
603 scope of this document. In performing registration techniques, either images aligned with the parametric  
604 maps or correlative images upon which the segmentation was performed are used as the target image for  
605 registration. The registered images are then interpolated from the source images. In interpolating  
606 parameter maps to match correlative images, tri-linear techniques are favored to avoid artifacts that may  
607 be associated with more advanced interpolation techniques.

608

609 **(c) Extraction of values for statistical comparison**

610

611 To derive values for statistical comparison from  $K^{\text{trans}}$  or  $\text{IAUGC}_{\text{BN}}$  parameter maps, median, mean and  
612 standard deviation of the pixel values should be calculated, and the median is considered the primary figure  
613 of merit. In a patient with multiple lesions due to metastatic disease, each lesion should be reported and  
614 tracked separately. In a patient with multiple lesions due to recurrent local tumor (for example, recurrent  
615 glioblastoma) per-patient figures of merit should be reported by aggregating the results of the multiple  
616 lesions.

617

618 **(d) Choice of time point for segmentation.**

619

620 As a rule, the  $K^{\text{trans}}$  or  $\text{IAUGC}_{\text{BN}}$  at a given time point should be extracted using tumor ROIs segmented from  
621 the same imaging examination. However, in the situation where anti-angiogenic therapies are evaluated  
622 and post-therapy imaging is performed within 72 hours of initial treatment with the anti-angiogenic agent,  
623 it is acceptable to use a recent (within 1 week) pre-therapy time point to provide the segmentation used to  
624 define the lesion on the immediate post-therapy imaging session. In this case, it is presumed that changes  
625 in the appearance of lesions on immediate post-therapy study are due to immediate decreases in  
626 permeability or blood flow rather than decrease in lesion volume.

627

628 In settings where analysis is performed retrospectively, all time points should be made available to the  
629 reader simultaneously to allow for consistency in choice of tumor(s) for segmentation, and to ensure that  
630 similar regions of large tumors have been sampled and segmented. In the case of manual VIF segmentation,  
631 such workflow analyses also allow for greater standardization of the region of the aorta or other artery  
632 used in the analysis. In such settings, the reader should be blinded to the nature of each time point, so that  
633 inherent bias in tumor and/or VIF segmentation does not influence the results.

## 634 6. Archival and Distribution of Data

635 Archival and data distribution procedures are recommended so that all analysis results can be recomputed  
636 for verification and validation purposes. In addition to saving of all original images in DICOM formats, the  
637 following information must be archived along with the image data:

### 638 6.1. Post-Processed Data

- 639 • **VIF:** Detailed specification of the vascular input function selection. This may include a  
640 binary mask of pixels selected for arterial input function, or may consist of a tabulated text  
641 file containing RAS coordinates co-ordinates of the VIF pixel locations.
- 642 • **Registration:** Recorded parameters and user inputs required for image registration, if used.  
643 Time-series image registration may be used to align data spatially over time. Any parameters  
644 which control the performance of the registration algorithm (metric used, optimization  
645 parameters, user click points/sub regions used for alignment, etc) must be stored in suitable  
646 format. It is preferred to save the registration transform parameters so that identical  
647 registration can be reproduced in a multi-center environment.

### 649 6.2. Analysis Results

650 All regions of interest where analysis is performed and statistics are computed should be saved. In addition,  
651 all computed maps ( $K^{\text{trans}}$  and  $\text{IAUGC}_{\text{BN}}$ ), should be saved in DICOM and DICOM secondary capture modes.  
652  $K^{\text{trans}} \text{ min}^{-1} * 10000$ .

### 653 6.3. Interpretation Results

654 All interpretation of results should be saved for purposes of verification and audit.

## 655 7. Quality Control

656 The following section deals with all aspects of quality control in DCE-MRI studies. This includes selecting  
657 and qualifying an MRI imaging center, MRI personnel, and specific MRI scanners. In addition, the use of  
658 phantom imaging (prior to study initiation and ongoing) is discussed. Finally, post image acquisition quality  
659 assessment is detailed. Details of these processes will vary for investigator-initiated single site studies  
660 versus sponsor-driven multi site studies.

661

662 Mechanisms for appropriate patient and tumor selection, image acquisition, and post processing are  
663 discussed throughout the document.

---

## 7.1. Selection of appropriate imaging centers for DCE-MRI studies

Typically sites are selected for DCE-MRI due to their competence in clinical oncology and access to a sufficiently large patient population under consideration. Sites must also be highly competent in clinical MRI techniques appropriate to the area(s) of anatomy to be imaged during the DCE-MRI study. In order to ensure high quality DCE-MRI results, it is essential to implement procedures that ensure quality assurance of the scanning equipment and reliable image acquisition methodology. These processes must be set-up at the outset, and followed throughout the duration of the study. A site “imaging capability assessment” prior to site selection is therefore a requirement for any DCE-MRI study. This will include assessment of:

- appropriate imaging equipment and quality control processes (see section 7.1.1)
- appropriate injector equipment and contrast media
- experienced MR technologists
- experienced MR radiologists
- experienced MR physicists or MR imaging scientists
- procedures to assure imaging protocol compliance during the trial

### 7.1.1 DCE-MRI Acquisition Scanner

DCE-MRI studies as developed in this profile require a 1.5 T MR scanner. The scanner software version should be identified and tracked across time, with updates and changes in scanner software noted during the course of a trial.

Proper coil maintenance must be performed to ensure adequate coil performance. It is beneficial to have alternate receiver coil systems available in the event that coil malfunction is identified prior to or during a DCE-MRI study.

The MRI scanner and receiver coils must undergo routine quality assurance and quality control processes (including preventive maintenance schedules) appropriate for clinical MRI applications. In addition, in order to assure adequate quantitative MR imaging results, additional quality control measures are required, as discussed below.

It is beneficial to identify and qualify more than one 1.5T MRI scanner at the site, if such are available for study use. This will ensure that if the primary MRI scanner is temporarily unavailable, the DCE-MRI study may proceed on a secondary scanner.

### 7.1.2 DCE-MRI Power Injector

A power injector is required for all DCE-MRI studies. The power injector needs to be properly serviced and calibrated.

### 7.1.3 MR Technologists or other Site Personnel performing DCE-MRI studies

MR technologists or other imaging expert(s) performing DCE-MRI procedures should be MR certified according to local regulations or institutional requirements. These individuals should have prior experience in conducting dynamic contrast enhanced imaging. The personnel should also be experienced in clinical

707 study related imaging and should be familiar with good clinical practices (GCP). Competence in the  
708 performance of DCE-MRI should never be limited to a single individual at the imaging center, as scheduled  
709 and unplanned personnel absences are to be expected in the course of a DCE-MRI trial.  
710

#### 711 **7.1.4 MR Radiologists or other anatomic experts**

712  
713 As tumor identification and selection is a critical component of the DCE-MRI study, sites performing DCE-  
714 MRI must have access to highly qualified MRI radiologists or other experts in MRI anatomic assessment.  
715 These individuals must be available during each DCE-MRI study to confirm adequate tumor selection and  
716 slab placement. In some settings, (e.g. brain tumors), it may be feasible for tumor identification and slab  
717 placement to be performed by the MR technologist, with oversight by a neuro-radiologist. In other cases  
718 (e.g. wide-spread metastatic disease in the chest, abdomen, or pelvis), it is accepted that a radiologist or  
719 other anatomic specialist must be available to identify tumor locations prior to contrast injection. It is  
720 expected that more than one anatomic specialist be available at a site performing the examination, should  
721 the primary anatomic specialists not be available for a given study.  
722

#### 723 **7.1.5 Site compliance with protocol requirements**

724  
725 Imaging centers participating in DCE-MRI trials must adhere to accepted standards of quality control in  
726 imaging studies. This includes processes to identify patients, who are participants in research studies,  
727 personnel familiar with local IRB and other regulatory practices, proper understanding of source  
728 documentation, and reporting of protocol deviations and adverse events. Imaging centers must be able to  
729 document their compliance with DCE-MRI procedures in order to facilitate central quality control and  
730 auditing processes. Centers participating in multi-site trials must be familiar with protocol-directed  
731 methods for image transfer of HIPAA-compliant anonymized imaging data, properly annotated, to central  
732 analytic laboratories.  
733  
734

## 735 **7.2 Site qualification process**

### 736 **7.2.1 Site readiness**

737  
738  
739 Site readiness for DCE-MRI should be documented prior to the initiation of the DCE-MRI trial. In single-site  
740 studies initiated by in-house investigators, imaging procedures should be reviewed with the DCE-MRI team  
741 prior to study initiation. In multi-site studies, site readiness assessment can begin with a simple  
742 questionnaire completed as a pre-qualification step. A subsequent site visit prior to DCE-MRI study  
743 initiation is recommended. During the site visit, study related imaging procedures and protocols are  
744 discussed. Ideally, all DCE-MRI scan parameters are reviewed and entered at the MR scanner at the time of  
745 the study visit. In some cases, initial phantom scanning can be performed during the site visit to familiarize  
746 local MR personnel with proper phantom handling, set-up, and scanning.  
747

### 748 **7.2.2 Scanner qualification**

749  
750 MR scanners should be identified based on their vendor, model, and machine name. Hardware  
751 specifications (maximum gradient strength, slew rate, etc.) should be documented. Software versions in

752 place at the time of trial initiation, and at all upgrades should be documented as well. Local receive coils to  
753 be used should be noted, with quality checks per local institutional methods documented. Power injector  
754 models should be noted, including date of most recent calibration.

### 755 **7.2.3 Phantom imaging**

756 To qualify the MRI scanner, a phantom imaging process is required. The QIBA DCE-MRI phantom, or a  
757 similar multi-compartment phantom with range of  $R_1$  relaxation rate values appropriate for DCE-MRI  
758 should be utilized. With the exceptions noted below, imaging of the phantom should otherwise be  
759 performed using the same  $R_1$  mapping ( $R_1 = 1/T_1$ ) and DCE-MRI acquisitions that are to be used in the  
760 clinical research protocol. Coil placement should approximate that which would be used by the site for the  
761 typical patient and anatomy of interest.  
762  
763

### 764 **7.2.4 Phantom imaging data analysis**

765 Phantom data should be analyzed in a uniform method by a centralized DCE-MRI image analysis center.  
766 Assurance should be made by the central site that the phantom scan orientation is correct, and appropriate  
767 image rotations or inversions were performed (and documented by the image analysis center).  
768

769 For all phantom image tests, a single central slice is utilized. Uniform 2cm ROI spheres are placed within  
770 each phantom compartment, avoiding the edges of the compartments where signal intensity may be  
771 altered by Gibbs lines or other artifacts. Mean and standard deviation of the signal intensities within each  
772 ROI should be noted. There are three categories of DCE-MRI phantom data analysis: signal stability, signal  
773 linearity, and  $R_1$  precision. In all cases, analysis should use a single central slice of the phantom data for  
774 analysis.  
775  
776

#### 777 **7.2.4.1. Signal stability**

778 The signal stability test is performed using the DCE-MRI acquisition method to be used for the dynamic  
779 gadolinium enhanced imaging. The duration of this scan should be at least 6 minutes to test magnet  
780 stability. A single  $R_1$  compartment with adequate SNR (10:1 or higher) is required. The mean SI in the ROI  
781 is then plotted over time. The plot should be linear and horizontal with no upward or downward trends.  
782 The root mean squared (rms) noise calculation should be similar across all aspects of the scan.  
783  
784

785 Marked deviations or drift of signal intensity over time indicate magnet instability, and should initiate a  
786 thorough evaluation of the magnet by the on-site MR physicist or site engineer prior to use in the DCE-MRI  
787 trial. The source of magnet instability should be determined and corrected prior to use in the DCE-MRI  
788 trial.  
789

#### 790 **7.2.4.2 Signal linearity**

791 In cases where signal intensity differences are to be used as a marker of tumor gadolinium concentration  
792 (see section 5), the linearity of MRI signal intensity with respect to  $R_1$  over a range of  $R_1$  values is required.  
793 While published guidelines on the allowed deviation from linearity do not exist, a linear correlation  
794 coefficient between SI and  $R_1$  of 0.9 or higher is expected.  
795  
796  
797

798 If a good linear correlation between SI and  $R_1$  is not achieved, it is recommended that the receive coil array  
799 used for phantom imaging be evaluated to ensure that coil failure was not a cause of the abnormal results.  
800 The phantom image may be repeated with a different local coil array, or with the body coil as receiver to  
801 further evaluate this issue.

802  
803 If linearity of SI vs.  $R_1$  is still not achieved, it is recommended that the phantom scan be repeated with a  
804 larger flip angle, in order to increase the relative  $T_1$  weighting of the images.

#### 805 806 *7.2.4.3 R1 precision*

807  
808 If  $T_1$ -dependent analysis is intended for the DCE-MRI study, the fidelity of  $R_1$  measurement should be  
809 assessed based on the phantom imaging. As uncertainty in the measurement of  $R_1$  is an important  
810 contributor to concentration measurement bias<sup>[48]</sup>, the measured phantom  $R_1$  values based on the VFA  
811 method (see Section 5) should be compared within the known  $R_1$  values calibrated based on non-flip angle  
812 dependent methods (such as IR-prepped imaging). Simulation studies suggest that variation in the  $R_1$  value  
813 by greater than 15% from actual may severely affect the reliability of the DCE-MRI quantification when  $R_1$  -  
814 dependent modeling of tumor gadolinium concentration in DCE-MRI studies is used.

815  
816 If accurate  $R_1$  values cannot be reproduced, it is recommended that  $R_1$  -dependent modeling not be  
817 performed.

#### 818 819 **7.2.5 Ongoing MRI scanner quality control**

820  
821 The phantom scans and analysis should be repeated at regular intervals, such as every 3 months, during the  
822 course of the study. Any changes to scanner equipment, including major hardware changes or any software  
823 version change, need to be documented and will result in the need for imaging qualification renewal prior  
824 to repeat imaging. In particular, it is strongly recommended that patients undergoing longitudinal study be  
825 scanned on the same MRI system with the same software version whenever possible. Sites performing  
826 DCE-MRI studies should be informed of planned software upgrades, when possible deferring such upgrades  
827 until serial imaging of all currently enrolled patients is complete.

### 828 829 **7.3. Quality Control of DCE-MRI studies**

#### 830 831 **7.3.1 Determination of suitable tumor lesions**

832  
833 Patients suitable for DCE-MRI analysis must possess at least one tumor  $\geq 2$ cm, well removed from areas  
834 subject to large degrees of cardiac pulsatility artifact, that is not largely cystic or necrotic. Determination of  
835 patient eligibility is usually based on pre-enrollment imaging (often CT or clinical MRI) which then serves as  
836 a baseline study for subsequent assessments for tumor response or progression. The site radiologist then  
837 reviews these images prior to enrollment to ascertain the location of the most suitable tumor lesion(s) for  
838 analysis.

#### 839 840 **7.3.2 Selection of target lesion**

841  
842 Once the MRI scan commences, the radiologist or anatomic expert will review the pre-gadolinium imaging  
843 to identify putative target lesions. The DCE-MRI study then proceeds with slab placement and  $T_1$

844 mapping/dynamic enhanced imaging once the target lesion is identified. Sites should strive to inspect  
845 these images to ensure absence of substantial artifacts (e.g., phase wrap, pulsatility) overlying the target  
846 lesion, with protocol specified adjustments to patient positioning and slab placement prior to continuing  
847 the DCE-MRI study. Once the final slab placement is confirmed, grid line overlays of the DCE-MRI slab on  
848 routine anatomic imaging (usually axial plane) is recommended to facilitate DCE-MRI slab placement on  
849 subsequent visits (e.g. by saving of a screen shot).

### 851 **7.3.3 Determination of subjects unsuitable for DCE-MRI analysis**

852  
853 Despite best efforts and protocol adherence, on occasion, a patient enrolled and imaged in DCE-MRI study  
854 will be found to be ineligible for subsequent analysis. Reasons for eliminating patients for analysis include:

- 855
- 856 • Lack of a tumor of suitable size in the usable DCE-MRI imaging volume
- 857 • Unacceptable pulsatility, wrap, or metallic artifact involving all tumors in the usable DCE-MRI  
858 imaging volume
- 859 • All target lesions in the DCE-MRI imaging volume determined to be largely cystic or necrotic
- 860 • Patients with significant amount of ascites since anti-angiogenic therapies can be very  
861 effective at reducing ascites and, hence, altering body weight, which may substantially affect  
862 the amount of gadolinium contrast agent administered.

863  
864 Determination of patient eligibility should be made by an independent reviewer who is blinded to other  
865 attributes of patient data, including (when applicable) randomization arm/drug treatment, toxicity, and  
866 clinical outcomes. Decisions on eligibility should be made on the basis of visual image assessment prior to  
867 analysis of DCE-MRI data. Quantitative criteria for defining tumors that are largely cystic or necrotic (such  
868 as percentage of pixels with enhancement above a certain threshold) should be defined in the protocol to  
869 avoid bias in decisions to eliminate patients from further DCE-MRI assessment.

### 871 **7.3.4 Determination of DCE-MRI exams unsuitable for DCE-MRI analysis**

872  
873 In addition, individual DCE-MRI examinations may be deemed nonanalyzable based on a variety of technical  
874 deviations. These may include:

- 875
- 876 • Failure of gadolinium injection
- 877 • Gross patient motion not correctable with motion correcting algorithms
- 878 • Failure of the imaging site to replicate the imaging parameters within acceptable standards  
879 of deviation from protocol specifications
- 880 • Failure of the imaging site to replicate anatomic DCE-MRI slab placement

881  
882 Whenever possible, all anticipated instances where individual DCE-MRI data will be removed from analysis  
883 should be prespecified in the DCE-MRI protocol.

### 885 **7.3.5 Editing of DCE-MRI exams prior to DCE-MRI analysis**

886  
887 It is recognized that DCE-MRI analysis requires post-processing of the DCE-MRI image sets. Most  
888 frequently, data sets will be subject to automated or semi-automated motion compensation schemes to  
889 eliminate or minimize the effects of image motion of subsequent DCE-MRI analysis. The methodology used

890 for such post processing should be documented, ideally in the DCE-MRI protocol or the standard operating  
891 procedures of the central analysis laboratory. Motion correction matrices keyed to each temporal phase  
892 may be documented as part of the analysis routine, in order to facilitate replication of the data analysis  
893 when required.

894  
895 In the course of post processing, individual phases of the DCE-MRI exam may be found to be severely  
896 compromised by image blur or degraded by other artifacts (such as random noise spikes). Judicious  
897 selection of phases to be eliminated for analysis may be made by the central analysis team, provided that  
898 the decision to eliminate such phases is determined prior to data analysis. Elimination of baseline or early  
899 post gadolinium phases is discouraged as such post processing may substantially alter the subsequent  
900 analysis. Data documenting these forms of post-processing should be maintained by the imaging analysis  
901 laboratory.

## 904 **8. Imaging-associated Risks and Risk Management**

905 MR safety considerations are to be established individually at each institution according to each  
906 institutions' radiology departmental guidelines and institutional review board (IRB) considerations to  
907 include policy guidelines on the following:

- 908 (1) laboratory screening for renal dysfunction prior to gadolinium based contrast administration
- 909 (2) contrast administration in pregnant patients and in patients who are lactating
- 910 (3) policy on patients receiving gadolinium based agents who have a positive history of a previous adverse  
911 event or events to iodinated or gadolinium based contrast agents to include serious and non-serious  
912 adverse events. The American College of Radiology Manual on Contrast Media Version 7 2010 can serve as  
913 a referenced guideline for each institutional policy development. This manual reflects policy statements  
914 previously released by the Food and Drug Administration (FDA) in the United States and its counterpart in  
915 the European Union, The Committee for Medicinal Products for Human Use (CHMP).

## 916 **IV. Compliance**

917 Typically clinical sites are selected due to their competence in oncology and access to a sufficiently large  
918 patient population under consideration. For DCE-MRI use as quantitative imaging biomarker it is essential  
919 to put some effort into an imaging capability assessment prior to final site selection for a specific trial. For  
920 imaging it is important to consider the availability of:

- 921 • appropriate imaging equipment and quality control processes,
- 922 • appropriate injector equipment and contrast media,
- 923 • experienced MR technologists for the imaging procedure, and
- 924 • processes that assure imaging protocol compliant image generation at the correct point in  
925 time.

### 926 **Acquisition Scanner**

927 1.5 T MR machines with 55-70 cm bores need to be available. The scanner needs to be under quality  
928 assurance and quality control processes (including preventive maintenance schedules) appropriate for  
929 quantitative MR imaging applications, which may exceed the standard requirements for routine clinical  
930 imaging or for MR facility accreditation purposes. The scanner software version should be identified and



931 tracked across time. It might be beneficial to identify and qualify a second scanner at the site, if available. If  
932 this is done prior to the study start there will be no difficulties later on in case the first scanner is  
933 temporarily unavailable. Practically speaking sites are encouraged to perform longitudinal treatment trials  
934 on one instrument.

### 935 **Contrast Inject Device**

936 A power injector is required for DCE-MRI studies. It needs to be properly serviced and calibrated.

### 937 **Software Analysis**

938 When a site is performing parametric image analysis and interpretation, a DCE-MRI tool that complies with  
939 the Tofts' model should be utilized. In addition, for multi-institutional trials a central reading site is  
940 assumed.

### 941 **Performing Site**

942 MR technologists running DCE-MRI procedures should be MR certified according to local regulations. The  
943 technologists should have prior experience in conducting dynamic contrast enhanced imaging. The person  
944 should be experienced in clinical study related imaging and should be familiar with good clinical practices  
945 (GCP). A qualified backup person is needed that should fulfill the same requirements. Contact details for  
946 both technologists should be available in case of any questions.

### 948 **Imaging qualification process:**

949  
950 The above-mentioned details can be obtained using a simple questionnaire as a pre-qualification step. If  
951 appropriate equipment and personnel are available, a site visit is recommended. During the site visit, study  
952 related imaging protocols are discussed and, ideally, all scan parameters are entered at the MR scanner.

953  
954 To qualify the scanner, a phantom imaging process is strongly recommended. The QIBA DCE-MRI phantom,  
955 or a similar multi-compartment phantom with range of relaxation rate ( $T_1$ ) values appropriate for the DCE-  
956 MRI study to be performed, should be used if the Profile Claim given above is to be assured. Data should be  
957 acquired from the multi-compartment phantom using the same  $T_1$  mapping and DCE-MRI acquisitions that  
958 will be used in the proposed clinical application or clinical research protocol (see Section 6).

959  
960 The phantom scans should be repeated on a regular interval (e.g 3 months) during the course of the study.  
961 Ongoing image quality inspection on a per scan basis is essential. Any changes to scanner equipment,  
962 including major hardware changes or any software version change, need to be documented and will result  
963 in the need for imaging qualification renewal.

### 965 **Site Analysis qualification:**

966  
967 The data analysis procedures to be used in the DCE-MRI application should be used to analyze the  $T_1$   
968 mapping data and results compared to the known  $T_1$  values of the various compartments. As uncertainty in  
969 the measurement of  $T_1$  is an important contributor to concentration measurement bias<sup>[48]</sup>, the measured  
970 values should compare within 15 % of the known values over a  $T_1$  range of approximately 50-1000 ms. The  
971 DCE-MRI data obtained from the phantom should be analyzed to confirm the correct temporal resolution  
972 and to provide SNR measurements and signal intensity vs.  $T_1$  characteristics for the specific DCE-MRI

973 acquisition protocol.

974  
975 Significant variations in any of these parameters during the course of an ongoing longitudinal study can  
976 affect the resulting imaging biomarker determinations, in the case of this specific claim  $K^{trans}$  and  $IAUGC_{BN}$ ,  
977 and such changes can readily occur if there are major changes in the scanner hardware or software, e.g., an  
978 update to the pulse sequence used for the DCE-MRI and/or  $T_1$  measurements or to the gradient subsystem.  
979 All results shall be documented and, if they pass the established acceptance values, will constitute the site  
980 qualification documentation for the DCE-MRI procedure. This process ensures study specific training of the  
981 site personnel and needs to be documented and signed.  
982

983

## References

- 984 1. Tofts, P. S. et al. Estimating kinetic parameters from dynamic contrast-enhanced T(1)-weighted MRI of  
985 a diffusable tracer: standardized quantities and symbols. (1999). *J Magn Reson Imaging* **10**:223-232
- 986 2. Evelhoch, J. L. Key factors in the acquisition of contrast kinetic data for oncology. (1999). *J Magn*  
987 *Reson Imaging* **10**:254-259
- 988 3. Ah-See, M. L. et al. Early changes in functional dynamic magnetic resonance imaging predict for  
989 pathologic response to neoadjuvant chemotherapy in primary breast cancer. (2008). *Clin Cancer Res*  
990 **14**:6580-6589
- 991 4. Dreys, J. et al. Phase I clinical study of AZD2171, an oral vascular endothelial growth factor signaling  
992 inhibitor, in patients with advanced solid tumors. (2007). *J Clin Oncol* **25**:3045-3054
- 993 5. Esserman, L. et al. Utility of magnetic resonance imaging in the management of breast cancer:  
994 evidence for improved preoperative staging. (1999). *J Clin Oncol* **17**:110-119
- 995 6. Hawighorst, H. et al. Angiogenic activity of cervical carcinoma: assessment by functional magnetic  
996 resonance imaging-based parameters and a histomorphological approach in correlation with disease  
997 outcome. (1998). *Clin Cancer Res* **4**:2305-2312
- 998 7. Hylton, N. Dynamic contrast-enhanced magnetic resonance imaging as an imaging biomarker. (2006).  
999 *J Clin Oncol* **24**:3293-3298
- 1000 8. O'Connor, J. P. et al. DCE-MRI biomarkers in the clinical evaluation of antiangiogenic and vascular  
1001 disrupting agents. (2007). *Br J Cancer* **96**:189-195
- 1002 9. Rosen, M. A. and Schnall, M. D. Dynamic contrast-enhanced magnetic resonance imaging for  
1003 assessing tumor vascularity and vascular effects of targeted therapies in renal cell carcinoma. (2007).  
1004 *Clin Cancer Res* **13**:770s-776s
- 1005 10. Solin, L. J. et al. Relationship of breast magnetic resonance imaging to outcome after breast-  
1006 conservation treatment with radiation for women with early-stage invasive breast carcinoma or ductal  
1007 carcinoma in situ. (2008). *J Clin Oncol* **26**:386-391
- 1008 11. Zahra, M. A. et al. Dynamic contrast-enhanced MRI as a predictor of tumour response to radiotherapy.  
1009 (2007). *Lancet Oncol* **8**:63-74
- 1010 12. Leach, M. O. et al. Assessment of antiangiogenic and antivascular therapeutics using MRI:  
1011 recommendations for appropriate methodology for clinical trials. (2003). *Br J Radiol* **76 Spec No**  
1012 **1**:S87-91
- 1013 13. NCI Recommendations for MR measurement methods at 1.5 Tesla and endpoints for use in Phase 1/2a  
1014 trials of anti-cancer therapeutics affecting tumor vascular function. Dynamic contrast MRI (DCE-MRI)  
1015 guidelines resulted from the NCI CIP MR Workshop on Translational Research in Cancer. (2004). MR  
1016 Workshop on Translational Research
- 1017 14. Ashton, E. et al. Scan-rescan variability in perfusion assessment of tumors in MRI using both model  
1018 and data-derived arterial input functions. (2008). *J Magn Reson Imaging* **28**:791-796
- 1019 15. Dowlati, A. et al. Novel Phase I dose de-escalation design trial to determine the biological modulatory  
1020 dose of the antiangiogenic agent SU5416. (2005). *Clin Cancer Res* **11**:7938-7944
- 1021 16. Ferl, G. Z. et al. An automated method for nonparametric kinetic analysis of clinical DCE-MRI data:  
1022 application to glioblastoma treated with bevacizumab. (2010). *Magn Reson Med* **63**:1366-1375
- 1023 17. Flaherty, K. T. et al. Pilot study of DCE-MRI to predict progression-free survival with sorafenib  
1024 therapy in renal cell carcinoma. (2008). *Cancer Biol Ther* **7**:496-501
- 1025 18. Galbraith, S. M. et al. Reproducibility of dynamic contrast-enhanced MRI in human muscle and  
1026 tumours: comparison of quantitative and semi-quantitative analysis. (2002). *NMR Biomed* **15**:132-142
- 1027 19. Liu, G. et al. Dynamic contrast-enhanced magnetic resonance imaging as a pharmacodynamic measure  
1028 of response after acute dosing of AG-013736, an oral angiogenesis inhibitor, in patients with advanced  
1029 solid tumors: results from a phase I study. (2005). *J Clin Oncol* **23**:5464-5473
- 1030 20. Morgan, B. et al. Dynamic contrast-enhanced magnetic resonance imaging as a biomarker for the

- 1031 pharmacological response of PTK787/ZK 222584, an inhibitor of the vascular endothelial growth  
1032 factor receptor tyrosine kinases, in patients with advanced colorectal cancer and liver metastases:  
1033 results from two phase I studies. (2003). *J Clin Oncol* **21**:3955-3964
- 1034 21. Ng, C. S. et al. Reproducibility of perfusion parameters in dynamic contrast-enhanced MRI of lung and  
1035 liver tumors: effect on estimates of patient sample size in clinical trials and on individual patient  
1036 responses. (2010). *AJR Am J Roentgenol* **194**:W134-40
- 1037 22. Padhani, A. R. et al. Reproducibility of quantitative dynamic MRI of normal human tissues. (2002).  
1038 *NMR Biomed* **15**:143-153
- 1039 23. Roberts, C. et al. Comparative study into the robustness of compartmental modeling and model-free  
1040 analysis in DCE-MRI studies. (2006). *J Magn Reson Imaging* **23**:554-563
- 1041 24. Stevenson, J. P. et al. Phase I trial of the antivascular agent combretastatin A4 phosphate on a 5-day  
1042 schedule to patients with cancer: magnetic resonance imaging evidence for altered tumor blood flow.  
1043 (2003). *J Clin Oncol* **21**:4428-4438
- 1044 25. Wedam, S. B. et al. Antiangiogenic and antitumor effects of bevacizumab in patients with  
1045 inflammatory and locally advanced breast cancer. (2006). *J Clin Oncol* **24**:769-777
- 1046 26. Zweifel, M. and Padhani, A. R. Perfusion MRI in the early clinical development of antivascular drugs:  
1047 decorations or decision making tools? (2010). *Eur J Nucl Med Mol Imaging* **37 Suppl 1**:S164-82
- 1048 27. Wang, Y., Spiller, M. and Caravan, P. Evidence for weak protein binding of commercial extracellular  
1049 gadolinium contrast agents. (2010). *Magn Reson Med* **63**:609-616
- 1050 28. Ledermann, H. P. et al. Screening for renal insufficiency following ESUR (European Society of  
1051 Urogenital Radiology) guidelines with on-site creatinine measurements in an outpatient setting. (2010).  
1052 *Eur Radiol* **20**:1926-1933
- 1053 29. Lee, Y. J. et al. Suboptimal contrast-enhanced carotid MR angiography from the left brachiocephalic  
1054 venous stasis. (1999). *J Magn Reson Imaging* **10**:503-509
- 1055 30. Tseng, Y. C. et al. Venous reflux on carotid computed tomography angiography: relationship with left-  
1056 arm injection. (2007). *J Comput Assist Tomogr* **31**:360-364
- 1057 31. You, S. Y. et al. Effects of right- versus left-arm injections of contrast material on computed  
1058 tomography of the head and neck. (2007). *J Comput Assist Tomogr* **31**:677-681
- 1059 32. B, H. and B, S. Determining optical flow. (1981). *Artificial intelligence*
- 1060 33. Sharp, G. C. et al. GPU-based streaming architectures for fast cone-beam CT image reconstruction and  
1061 demons deformable registration. (2007). *Phys Med Biol* **52**:5771-5783
- 1062 34. Pluim, J. P., Maintz, J. B. and Viergever, M. A. Mutual-information-based registration of medical  
1063 images: a survey. (2003). *IEEE Trans Med Imaging* **22**:986-1004
- 1064 35. Parker, G. J. et al. Probing tumor microvasculature by measurement, analysis and display of contrast  
1065 agent uptake kinetics. (1997). *J Magn Reson Imaging* **7**:564-574
- 1066 36. Vonken, E. J. et al. Measurement of cerebral perfusion with dual-echo multi-slice quantitative dynamic  
1067 susceptibility contrast MRI. (1999). *J Magn Reson Imaging* **10**:109-117
- 1068 37. Ashton, E., McShane, T. and Evelhoch, J. Inter-operator variability in perfusion assessment of tumors  
1069 in MRI using automated AIF detection. (2005). *Med Image Comput Comput Assist Interv* **8**:451-458
- 1070 38. Rijpkema, M. et al. Method for quantitative mapping of dynamic MRI contrast agent uptake in human  
1071 tumors. (2001). *J Magn Reson Imaging* **14**:457-463
- 1072 39. Tofts, P. S. and Kermode, A. G. Measurement of the blood-brain barrier permeability and leakage  
1073 space using dynamic MR imaging. 1. Fundamental concepts. (1991). *Magn Reson Med* **17**:357-367
- 1074 40. E, A. System and method for Identifying Optimized Blood Signal in Medical Images to Eliminate Flow  
1075 Artifacts. (2007).
- 1076 41. McGrath, D. M. et al. Comparison of model-based arterial input functions for dynamic contrast-  
1077 enhanced MRI in tumor bearing rats. (2009). *Magn Reson Med* **61**:1173-1184
- 1078 42. Meng, R. et al. Comparison between population average and experimentally measured arterial input  
1079 function in predicting biopsy results in prostate cancer. (2010). *Acad Radiol* **17**:520-525

- 
- 1080 43. Parker, G. J. et al. Experimentally-derived functional form for a population-averaged high-temporal-  
1081 resolution arterial input function for dynamic contrast-enhanced MRI. (2006). *Magn Reson Med*  
1082 **56**:993-1000
- 1083 44. Wang, Y. et al. Feasibility of using limited-population-based arterial input function for  
1084 pharmacokinetic modeling of osteosarcoma dynamic contrast-enhanced MRI data. (2008). *Magn Reson*  
1085 *Med* **59**:1183-1189
- 1086 45. Mouridsen K et al. Subject-specific AIF optimizes reproducibility of perfusion parameters in  
1087 longitudinal DSC-MRI in comparison to session and population level AIF. (2011). *International*  
1088 *Society of Magnetic Resonance in Medicine*
- 1089 46. S, K. Peripheral blood flow measurement. (1960). *Methods in Medical research* **8**
- 1090 47. Ahearn, T. S. et al. The use of the Levenberg-Marquardt curve-fitting algorithm in pharmacokinetic  
1091 modelling of DCE-MRI data. (2005). *Phys Med Biol* **50**:N85-92
- 1092 48. Schabel, M. C. and Parker, D. L. Uncertainty and bias in contrast concentration measurements using  
1093 spoiled gradient echo pulse sequences. (2008). *Phys Med Biol* **53**:2345-2373
- 1094 49. Oken, M. M. et al. Toxicity and response criteria of the Eastern Cooperative Oncology Group. (1982).  
1095 *Am J Clin Oncol* **5**:649-655
- 1096
- 1097

---

## 1098 Appendices

### 1099 Appendix A: Acknowledgements and Attributions

1100	<i>I. Executive Summary</i>	<i>Jeffrey Evelhoch</i>
1101	<i>II. Clinical Context and Claims</i>	<i>Mitchell Schnall</i>
1102	<i>III. Profile Details</i>	
1103	1. <i>Subject Handling</i>	<i>Alex Guimaraes</i>
1104	2. <i>Imaging Procedure</i>	<i>Ed Jackson/Sandeep Gupta</i>
1105	3. <i>Image Post-processing</i>	<i>Sandeep Gupta</i>
1106	4. <i>Parametric image formation</i>	<i>Ed Ashton</i>
1107	5. <i>Parametric image analysis</i>	<i>Dan Barboriak</i>
1108	6. <i>Archival and Distribution of Data</i>	<i>Sandeep Gupta</i>
1109	7. <i>Quality Control</i>	<i>Mark Rosen</i>
1110	8. <i>Imaging associated Risks and Risk Management</i>	<i>Orest Boyko</i>

1111

---

## 1112 **Appendix B: Conventions and Definitions**

### 1113 **B.1 List of Abbreviations**

- 1114
- 1115 - VIF: Vascular input function
- 1116 - DCE-MRI: Dynamic contrast enhanced magnetic resonance imaging
- 1117 - ECOG: Eastern Cooperative Oncology Group
- 1118 - eGFR: estimated Glomerular Filtration Rate
- 1119 - Gd-DTPA: Gadolinium – diethylene triamine pentaacetic acid
- 1120 - IAUGCBN: Initial area under the Gadolinium concentration blood normalized
- 1121 - Ktrans: Permeability transfer constant
- 1122 - QIBA: Quantitative Imaging Biomarkers Alliance
- 1123 - ROI: Region of Interest
- 1124 - VEGF: Vascular Endothelial Growth Factor
- 1125 - VFA: Variable Flip angle
- 1126 - GCP: Good Clinical Practice
- 1127 - SPGR (Spoiled Gradient Recalled)

### 1128 **B.2 ECOG Performance Status Descriptions, by grade:** <sup>[49]</sup>

- 1129
- 1130
- 1131 0: Fully active, able to carry on all pre-disease performance without restriction
- 1132 1: Restricted in physically strenuous activity but ambulatory and able to carry out work of a light or
- 1133 sedentary nature, e.g., light-house work, office work
- 1134 2: Ambulatory and capable of all self-care but unable to carry out any work activities. Up and about more
- 1135 than 50% of waking hours
- 1136 3: Capable of only limited self-care, confined to bed or chair more than 50% of waking hours
- 1137 4: Completely disabled. Cannot carry on any self-care. Totally confined to bed or chair
- 1138 5: Dead
- 1139

1140 **Appendix C: Spreadsheet on reproducibility data**

Reference	Year	Field Strength (T)	Organ System	N	Contrast Agent	Injection Rate	Flush	Temporal Resolution (s) / # sections
Ng, Raunig, Jackson, et al	2010	1.5	Liver / Lung	12 (lung) / 11 (liver)	Magnevist (0.1 mmol/kg)	3 ml/s	20 ml saline @ 3 ml/s	10.4 / 10
Ferl, Lu, Friesenhahn, et al	2010	1.5	Brain (GBM)	16	Magnevist (0.1 mmol/kg)	3 ml/s	Not stated	4.8 / 16
Ashton, Raunig, Ng, et al	2008	1.5	Liver / Lung	12 (lung) / 12 (liver)	Magnevist (0.1 mmol/kg)	3 ml/s	20 ml saline @ 3 ml/s	10.4 / 10
Lankester, Taylor, Stirling, et al	2007	1.5	Various tumors (pelvic)	20	Magnevist (0.1 mmol/kg)	4 ml/s	Not stated	12.0 / 4
Roberts, Issa, Stone, et al	2006	1.5	Brain and Abdomen	4 (brain) / 9 (abdo)	Omniscan (brain); Magnevist (abdo); 0.1 mmol/kg	Hand injected (3-4 s)	Brain: same volume; Abdo: not stated	8s / 25 (brain); 8 s early and 75 s late (abdo)
Morgan, Utting, Higginson, et al	2006	1.5	Various tumors (liver, lung, lymph node)	10	Magnevist or Omniscan (0.1 mmol/kg) or	Manually, less than 5 s	Not stated	0.5 / 1
Lankester, Taylor, Stirling, et al	2005	1.5	Various tumors (body)	20	Magnevist (0.1 mmol/kg)	4 ml/s	Not stated	Not stated
Jackson, Jayson, Li, et al.	2003	1.5	Brain (glioma)	9	Omniscan (0.1 mmol/kg)	Hand injected (3-4 s)	Saline at same volume and injection duration	5.1 - 8.7 / 24
Galbraith, Lodge, Taylor et al	2002	1.5	Various tumors (body)	16	Magnevist (0.1 mmol/kg)	Not stated	Not stated	11.9
Rijpkema, Kaanders, Joosten et al	2001	1.5	Various (6 H&N; 2 brain; 3 prostate)	11	Magnevist (15 ml)	2.5 ml/s	Not stated	2

1:  
1142  
1143



Model (Tofts, GKM, etc)	Whole ROI or Pixelwise?	Parameters Reported	AIF	T1 Correction?	If yes, T1 mapping technique?	Fitted Data Type ( $\Delta[Gd]$ , $\Delta SI$ , $\Delta SI/S_0$ )
2 param GKM	Pixel	Ktrans, kep, IAUC90 <sub>BN</sub>	Yes, automated	No		SI
Deconvolution and 3-param GKM	Pixel	Ktrans, ve	Yes (venous)	Yes	VFA (5, 10, 15, 20, 25, 30)	[Gd]
2 param GKM	Pixel	Ktrans, kep, IAUC90 <sub>BN</sub>	Yes, automated	No		SI
IAUGC, Kety (=Tofts?)	Pixel	IAUGC60, Ktrans, kep, Ve	No (pooled data)	Yes	Proton density reference	[Gd]
IAUC, Tofts (2 param), Tofts (3 param)	Pixel	IAUC60 (Model 1); Ktrans, ve (Model 2); Ktrans, ve, vp (Model 3)	No (Model based)	Yes	VFA (2, 20, 35: brain; 2, 13, 28: abdo)	[Gd]
IAUC, Tofts (2 compart)	Not specified	PE, IAUC60, IAUC180, Ktrans	Yes and No, local data, and modified on published data	Yes?	IR	
IAUC, Tofts (2 param)	Pixel	IAUCGC60, Ktrans	No (Model based)	Yes	Proton density reference	[Gd]
2 param GKM	Pixel	Ktrans, ve	Yes (sagittal sinus, fitted to biexponential)	Yes	VFA (2, 10, 35)	[Gd]
IAUC, Tofts (2 param)	Pixel	IAUC90, Ktrans, kep, ve	No (Model based)	Yes	Proton density reference	[Gd]
3 param GKM	Pixel	kep	Yes	Yes	Proton density reference	[Gd]

1144

1145

Motion Correction?	Primary Findings (test/retest CV, CI, etc)	Additional Findings	Reference
Yes	Within Patient CV. Ktrans: liver:8.9%, lung:17.9%; IAUC: liver:9.9%, lung:18.2%.	Sample size requirements of liver and lung for %change in Ktrans and IAUC	Ng, Raunig, Jackson, et al
None stated	Repeat baseline CV%. Ktrans: 13.6%, ve: 23.6%	Deconvolution method: AUC/MRT: 10.7%, AUC: 12.7%	Ferl, Lu, Friesenhahn, et al
Yes	Within Patient CV. Ktrans: liver:10.6%, lung:19.3%; IAUC: liver:9.8%, lung:15.7%.	Also used Tofts model derived method; Within Patient CV (Ktrans, kep). Ktrans: liver:35.6%, lung:20.7%; IAUC: liver:33.1%, lung:18.9%.	Ashton, Raunig, Ng, et al
None stated	Within Patient CV. Ktrans: 20.3%, Ve: 8.3%, kep: 17.4%, IAUGC: 12.1%	Additional results to previous paper of 2005	Lankester, Taylor, Stirling, et al
None stated	RMS CV%. IAUC60: 19%; Model 2: Ktrans:13%, ve:11%; Model 3: Ktrans:19%, ve:14%, vp:30%	Ktrans vs IAUC60 correlation	Roberts, Issa, Stone, et al
Yes	Within Patient CV. Ktrans: 19.1%, IAUC60: 15.8%, IAUC180: 16.1%, PE: 15.9%	Correlation of IAUC60 and IAUC180 with Ktrans after treatment	Morgan, Utting, Higginson, et al
None stated	Within Patient CV. Ktrans: 20.3%, IAUGC: 12.1%		Lankester, Taylor, Stirling, et al
None stated	Within Patient CV. Ktrans: 7.7%; ve: 6.2%	Max intensity change / unit time (MITR): 17.9%; Time to 90% enhancement (T90): 7.1%; Tumor volume: 4.0%; Native tumor T1 relaxation rate (R <sub>1,0</sub> ): 9.2%	Jackson, Jayson, Li, et al.
None stated	Within Patient CV. Ktrans: 24%, kep: 21%, ve: 8.5%, IAUC90:12%	Muscle data (whole ROI only); Whole ROI tumor data	Galbraith, Lodge, Taylor et al
None stated	No statistical difference in kep in 10 of 11 patients (Student's t-test, p:0.05)		Rijpkema, Kaanders, Joosten et al

1146

1147

1148

---

**1149 Appendix D: Model-specific Instructions and Parameters**

1150 The presence of specific product models/versions in the following tables should not be taken to imply that  
1151 those products are fully compliant with the QIBA Profile. Compliance with a profile involves meeting a  
1152 variety of requirements of which operating by these parameters is just one. To determine if a product (and  
1153 a specific model/version of that product) is compliant, please refer to the QIBA Conformance Document for  
1154 that product. G.1. Image Acquisition Parameters The following technique tables' list acquisition parameter  
1155 values for specific models/versions that can be expected to produce data meeting the requirements of  
1156 Section 7.1.

1157 These technique tables may have been prepared by the submitter of this imaging protocol document, the  
1158 clinical trial organizer, the vendor of the equipment, and/or some other source. (Consequently, a given  
1159 model/version may appear in more than one table.) The source is listed at the top of each table.

1160 Sites using models listed here are encouraged to consider using these parameters for both simplicity and  
1161 consistency. Sites using models not listed here may be able to devise their own acquisition parameters that  
1162 result in data meeting the requirements of Section 7.1 and conform to the considerations in Section 13.

1163 In some cases, parameter sets may be available as an electronic file for direct implementation on the  
1164 imaging platform.

1165

1166 **Siemens**  
 1167 QIBA DCE-MRI Abdominal Protocol for VA30 Software  
 1168  
 1169

parameter	value	notes
Routine tab		
slabs	1	
distance factor	irrevelant	
position	as needed	
orientation	coronal	
phase enc. dir.	R >> L	
rotation	0.0 deg	
phase oversampling	0%	
slice oversampling	0%	
slices per slab	26	Reconstructed images, interpolated by zero-filling. The slab thickness is 4.25 x 26 = 110.5 mm
FoV read	400	
FoV phase	81.3%	325 mm
slice thickness	4.25 mm	For 3-D, this is the slice <i>spacing</i> . The true slice thickness is this number divided by the slice resolution, in this case $4.25 / 0.62 = 6.85$ mm.
TR	5.03 ms	
TE	1.9 ms	
averages	1	NEX
concatenations	1	
filter	none	
coil elements	as needed	
Contrast tab		
flip angle	30 deg	

fat suppression	none	
water supp.	none	
Dixon	no	
save original images	on	
averaging mode	short term	
reconstruction	magnitude	
measurements	40	
measurement series	each measurement	
pause after measurement	0 sec	
Resolution tab		
base resolution	256	readout pixel size 1.56 mm
phase resolution	62%	phase pixel size 2.52 mm
slice resolution	62%	Controls zero-filling in slice. If no partial Fourier processing is used, 16 partitions are acquired. The raw matrix is padded with 10 zeros to reconstruct 26 slices: $16 / 0.62 = 26$ . Divide the slice spacing by the slice resolution to get the slice thickness: $4.25 / 0.62 = 6.85$ mm
phase partial Fourier	choose 7/8ths here or below (slice)	If 7/8ths is chosen, partial Fourier processing is used to reduce the number of acquired lines to: $256 \times 0.62 \times 0.813 \times 7/8 = 113$
slice partial Fourier	choose 7/8ths here or above (phase)	If 7/8ths is chosen, 14 partitions are acquired to provide the data for 16. Ten additional zeros are added to reconstruct 26 slices.
interpolation	on	In-plane zero-filling to 512 x 512.
PAT mode	none	No SENSE or GRAPPA
matrix coil mode	as needed	
image filter	off	
distortion correction	off	also called "large FoV filter"
prescan normalize	off	
normalize	off	Acts on individual slices, so must be turned off.
raw filter	off	
elliptical filter	off	

Geometry card		
multi-slice mode	irrelevant	
series	irrelevant	
special sat.	none	
(remainder)		May be ignored.
System Card		
shim mode	standard	
save uncombined	off	
adjust with body coil	off	
Physio card		
1 <sup>st</sup> signal/mode	none	
rsp. control	off	
Inline card		
3D centric reordering	off	
(remainder)	off	
Sequence card		
introduction	off	
dimension	3D	
elliptical scanning	off	
asymmetric echo	allowed, weak	
contrasts	1	
bandwidth	250 Hz/pixel	Corresponds to $\pm 32$ KHz.

optimization	min TE	
RF pulse type	normal	
gradient mode	fast	
excitation	slab-sel.	
RF spoiling	on	For the FLASH sequence.
Tool Tips		Roll the cursor over the appropriate item to view these.
readout echo position	38%	Roll over “echo asymmetry.”
matrix size	129 x 256	Roll over “phase resolution.” This size includes the effects of reduced pixel resolution and rectangular FoV.
slab thickness	110 mm	
pulse sequence	fl3d_vibe	Roll over the pulse sequence abbreviation.

1170

1171 SNR protocol: change measurements to 8 and flip angle to 15°.

1172 Variable flip angle protocol for T<sub>1</sub>: one measurement, 4 averages, and flip angles of 2°, 5°, 10°, 15°, 20°,

1173 25°, and 30°.

1174  
1175  
1176  
1177

QIBA DCE-MRI Abdominal Protocol for VB15, VB17, and VD11 Software  
These are the 400 Hz/pixel protocols.

parameter	value	notes
Routine tab		
slabs	1	
distance factor	irrelevant	
position	as needed	
orientation	coronal	
phase enc. dir.	R >> L	
rotation	0.0 deg	
phase oversampling	0%	
slice oversampling	0%	
slices per slab	26	Reconstructed images, interpolated by zero-filling. The slab thickness is $4.25 \times 26 = 110.5$ mm
FoV read	400	
FoV phase	81.3%	325 mm
slice thickness	4.25 mm	For 3-D, this is the slice <i>spacing</i> . The true slice thickness is this number divided by the slice resolution, in this case, $4.25 / 0.62 = 6.85$ mm.
TR	3.61 ms 3.91 ms 4.76 ms	VD11, Aera VB17, Espree VB15B, Verio
TE	1.49 ms 1.48 ms 1.43 ms	VD11, Aera VB17, Espree VB15B, Verio
averages	1	NEX
concatenations	1	
filter	none	
coil elements	as needed	
Contrast tab		



flip angle	30 deg	
fat suppression	none	
water suppression	none	
Dixon	no	
save original images	on	
averaging mode	short term	
reconstruction	magnitude	
measurements	50	as needed
measurement series	each measurement	
pause after measurement	0 sec	for all measurements
Resolution tab		
base resolution	256	readout pixel size 1.56 mm
phase resolution	62%	phase pixel size 2.52 mm
slice resolution	62%	Controls zero-filling in slice. Sixteen partitions are acquired. The raw matrix is padded with 10 zeros to reconstruct 26 slices: $16 / 0.62 = 26$ Divide the slice spacing by the slice resolution to get the slice thickness: $4.25 / 0.62 = 6.85$ mm
phase partial Fourier	off	No further reduction in the number of acquired lines: $256 \times 0.62 \times 0.813 = 129$
slice partial Fourier	off	No further reduction in the number of acquired partitions (16).
interpolation	on	In-plane zero-filling to 512 x 512.
PAT mode	none	No SENSE or GRAPPA
matrix coil mode	as needed	
image filter	off	
distortion correction	off	
prescan normalize	off	
normalize	off	Acts on individual slices, so must be turned off.
<b>B</b> <sub>1</sub> filter	off	
raw filter	off	
elliptical filter	off	

POCS	off	
Geometry card		
multi-slice mode	irrelevant	
series	irrelevant	
special sat.	none	
Set-n-Go Protocol	off	
inline composing	off	
System Card		
shim mode	tune up	
save uncombined	off	
adjust with body coil	off	
confirm freq. adjustment	off	
Physio card		
1 <sup>st</sup> signal/mode	none	
resp. control	off	
Inline card		
3D centric reordering	off	
(remainder)	off	
Sequence card		
introduction	off	
dimension	3D	
elliptical scanning	off	

asymmetric echo	allowed, weak	
contrasts	1	
bandwidth	400 Hz/pixel	Corresponds to $\pm 51.2$ KHz.
optimization	min TE	
RF pulse type	normal	
gradient mode	fast normal fast	VD11, Aera VB17, Espree VB15B, Verio
excitation	slab-sel.	
RF spoiling	on	For the FLASH sequence.
Tool Tips		Roll the cursor over the appropriate item to view these.
readout echo position	38%	Roll over "echo asymmetry."
matrix size	129 x 256	Roll over "phase resolution." This size includes the effects of reduced pixel resolution and rectangular FoV.
slab thickness	110 mm	
pulse sequence	fl3d_vibe	Roll over the pulse sequence abbreviation.

1178

1179

1180 SNR protocol: change measurements to 8 and flip angle to 15°.

1181 Variable flip angle protocol for T<sub>1</sub>: one measurement, 4 averages, and flip angles of 2°, 5°, 10°, 15°, 20°,  
1182 25°, and 30°.

1183

1184

**GE**

<b>GE Scanners</b>	
<b>DCE Scan</b>	
B0:	1.5T
Grad Subsystem:	BRM, TRM (Zoom), CRM, XRM
Coil:	Torso Phased Array
Sequence:	3D FSPGR
Slice orientation:	Oblique Coronal
Imaging Options:	EDR, MPH, ZIP2, ZIP512
User CVs:	Turbo <sup>1</sup> =2 / Slice res=100%
TE (ms):	0.9
TR (ms):	4.1 <sup>2</sup>
Flip Angle (deg):	30
Bandwidth:	+/- 32 kHz
NEX:	1
FOV (cm):	40
Phase FOV:	0.8
Slice Thickness (mm):	5
# locs per slab:	16
Acquisition matrix:	256 x 160
Freq Direction:	S/I
Scan time/volume:	8.5 <sup>2</sup> sec
Scan time / 40 volumes:	5:40 <sup>2</sup> min
<b>T1 Mapping Protocol</b>	
B0:	1.5T
Grad Subsystem:	BRM, TRM (Zoom), CRM, XRM
Coil:	Torso Phased Array
Sequence:	3D FSPGR
Slice orientation:	Oblique Coronal
Imaging Options:	EDR, MPH, ZIP2, ZIP512
User CVs:	Turbo <sup>1</sup> =0 / Slice res=100%
TE (ms):	1.0
TR (ms):	5.2 <sup>2</sup>
Flip Angle (deg):	2, 5, 10, 15, 20, 25, 30
Bandwidth:	+/- 32 kHz
NEX:	4
FOV (cm):	40
Phase FOV:	0.8
Slice Thickness (mm):	5
# locs per slab:	16
Acquisition matrix:	256 x 160
Freq Direction:	S/I
Acq Time (min):	43 <sup>2</sup> sec / flip angle

**Notes:**

1. Turbo (User CV or Advanced) should be set to 2 (fastest) for the DCE scan, but should be set to 0 (slowest) for the T1 mapping scans. If Turbo is set to 2 for the T1 mapping scans, the value of TE will change with flip angle, particularly for larger flip angle values.
2. The value of TR and, therefore, the scan time/volume and total scan time, will change slightly depending on the particular gradient subsystem used for the scans. The values above were obtained on a CRM platform and similar or slightly longer values can be obtained on BRM platforms, TRM platforms (if in Zoom Mode; substantially longer TR values are obtained if in Whole Mode), and XRM platforms.

1185  
1186  
1187  
1188  
1189  
1190  
1191  
1192  
1193

1194

**Phillips**

INFO PAGE		GEOMETRY		CONTRAST	
Total scan duration	05:50.3	Nucleus	H1	Scan type	Imaging
Rel. signal level (%)	100	Coil selection	SENSE-XL-Torso	Scan mode	3D
Act. TR/TE (ms)	5.0 / 2.4	element selection	All	technique	FFE
Dyn. scan time	00:08.3	connection	d	Contrast enhancement	T1
Time to k0	00:01.9	Dual coil	no	Acquisition mode	cartesian
ACQ matrix M x P	256 x 162	CLEAR	no	Fast Imaging mode	none
ACQ voxel MPS (mm)	1.64 / 2.10 / 4.00	FOV FH (mm)	420	3D non-selective	no
REC voxel MPS (mm)	0.82 / 0.82 / 2.00	RL (mm)	341.25	Echoes	1
Scan percentage (%)	78.125	AP (mm)	48	partial echo	no
Act. WFS (pix) / BW (Hz)	0.692 / 313.8	Voxel size FH (mm)	1.64	shifted echo	no
Min. WFS (pix) / Max. BW (Hz)	0.148 / 1464.8	RL (mm)	2.1	TE	shortest
SAR / whole body	< 40 % / 1.6 W/kg	AP (mm)	2	Flip angle (deg)	30
Whole body / level	< 1.6 W/kg / normal	Recon voxel size (mm)	0.8203125	TR	shortest
B1 rms [uT]	2.865556	Fold-over suppression	yes	Halfscan	yes
PNS / level	44 % / normal	Slice oversampling	user defined	factor Y	0.65
Sound Pressure Level (dB)	20.09241	oversample factor	1	factor Z	0.85
<b>MOTION</b>		Reconstruction matrix	512	Water-fat shift	maximum
Cardiac synchronization	no	SENSE	no	Shim	volume
Respiratory compensation	no	k-t BLAST	no	ShimAlign	no
Navigator respiratory comp	no	Overcontiguous slices	yes	Fat suppression	no
Flow compensation	no	Stacks	1	Water suppression	no
fMRI echo stabilisation	no	slices	24	MTC	no
NSA	2	slice orientation	coronal	Research prepulse	no
SMART	yes	fold-over direction	RL	Diffusion mode	no
<b>DYN/ANG</b>		fat shift direction	F	Elastography mode	no
Angio / Contrast enh.	contrast enh.	Chunks	1	SAR mode	high
Quantitative flow	no	PlanAlign	no	B1 mode	default
CE profile order	linear	REST slabs	0	PNS mode	high
Manual start	no	Catheter tracking	no	Gradient mode	maximum
Dynamic study	individual	Interactive positioning	no	SofTone mode	no
dyn scans	42	Allow table movement	no		
recon multiplier	1	<b>OFFC/ANG</b>			
dyn scan times	user defined	Stacks	1		
(mm:ss)	shortest (00:00.0), shortest (00:08.4), shortest (00:16.7), shortest (00:25.0), shortest (00:33.4), manual (00:41.7), shortest (00:50.1), shortest (00:58.4), shortest (01:06.8), shortest (01:15.1), shortest (01:23.4), shortest (01:31.8), shortest (01:40.1), shortest (01:48.5), shortest (01:56.8)	Stack Offc. AP (P=+mm)	0		
		RL (L=+mm)	0		
		FH (H=+mm)	0		
		Ang. AP (deg)	0		
		RL (deg)	0		
		FH (deg)	0		
		Shim Size AP (mm)	100		
		RL (mm)	100		
		FH (mm)	100		
		Offc. AP (P=+mm)	0		
		RL (L=+mm)	0		
		FH (H=+mm)	0		
		Ang. AP (deg)	0		

1195

1196

1197

	shortest (01:48.5), shortest (01:56.8), shortest (02:05.1), shortest (02:13.5), shortest (02:21.8), shortest (02:30.2), shortest (02:38.5), shortest (02:46.8), shortest (02:55.2), shortest (03:03.5), shortest (03:11.9), shortest (03:20.2), shortest (03:28.5), shortest (03:36.9), shortest (03:45.2), shortest (03:53.6), shortest (04:01.9), shortest (04:10.2), shortest (04:18.6), shortest (04:26.9), shortest (04:35.3), shortest (04:43.6), shortest (04:51.9), shortest (05:00.3), shortest (05:08.6), shortest (05:17.0), shortest (05:25.3), shortest (05:33.6), shortest (05:42.0), _ _ _ _ _ _ _ _ _ _ _ _ _ _ _		
		Ang. AP (deg)	0
		RL (deg)	0
		FH (deg)	0
dummy scans	0		
immediate subtraction	no		
fast next scan	no		
synch. ext. device	no		
prospect. motion corr.	no		
Keyhole	no		
Arterial Spin labeling	no		
<b>POST/PROC</b>			
Preparation phases	auto		
Manual Offset Freq.	no		
SmartPlan survey	no		
B0 field map/Dixon	no		
B1 field map	no		
MIP/MPR	no		
Images	M, no, no, no		
Autoview image	M		
Calculated images	no, no, no, no		
Reference tissue	Liver		
Preset window contrast	soft		
Reconstruction mode	real time		
reuse memory	no		
Save raw data	no		
Hardcopy protocol	no		

1198  
1199  
1200  
1201